

DIN EN ISO 15798:2022-05 (E)

Ophthalmic implants - Ophthalmic viscosurgical devices (ISO 15798:2022)

Contents		Page
	European foreword	3
	Foreword	4
1	Scope	6
2	Normative references	6
3	Terms and definitions	7
4	Intended performance	8
5	Design attributes	9
5.1	General.....	9
5.2	Characterization of the components.....	9
5.3	Characterization of the finished product.....	9
5.3.1	General.....	9
5.3.2	Absolute complex viscosity.....	10
5.3.3	Chemical and biological contaminants.....	10
5.3.4	Concentration.....	10
5.3.5	Elasticity.....	10
5.3.6	Molecular mass distribution.....	10
5.3.7	Osmolality.....	10
5.3.8	Particulates.....	11
5.3.9	pH.....	11
5.3.10	Refractive index.....	11
5.3.11	Shear viscosity.....	11
5.3.12	Spectral transmittance.....	11
5.4	Usability.....	12
6	Design evaluation	12
6.1	General.....	12
6.2	Evaluation of biological safety.....	12
6.2.1	General.....	12
6.2.2	Bacterial endotoxins test.....	13
6.2.3	Clearance of residual OVD from the anterior chamber.....	13
6.2.4	Degradation and toxicokinetics.....	13
6.2.5	Evaluation of inflammation and intraocular pressure.....	13
6.3	Clinical evaluation.....	13
6.3.1	General.....	13
6.3.2	Clinical investigation design.....	14
6.3.3	Corneal endothelial cell density.....	14
6.3.4	Postoperative inflammation.....	14
6.3.5	Post-operative intraocular pressure change.....	15
6.3.6	Adverse events.....	15
7	Sterilization	15
8	Product stability	16
9	Integrity and performance of the delivery system	16
10	Packaging	16
10.1	Protection from damage during storage and transport.....	16
10.2	Maintenance of sterility in transit.....	16
11	Information to be supplied by the manufacturer	16
	Annex A (normative) Intraocular implantation test	18
	Annex B (informative) Patient numbers for clinical investigation of intraocular pressure	21
	Annex C (informative) Analyses of OVD clinical data	22
	Bibliography	24