

# ISO/TR 22676:2006-10 (E)

## Prosthetics - Testing of ankle-foot devices and foot units - Guidance on the application of the test loading conditions of ISO 22675 and on the design of appropriate test equipment

---

### Contents

Page

Foreword .....	vi
Introduction .....	vii
<b>1</b> <b>Scope .....</b>	<b>1</b>
<b>2</b> <b>Guidance on the specification of the test loading conditions of ISO 22675 .....</b>	<b>1</b>
2.1 <b>General .....</b>	<b>1</b>
2.2 <b>Directions of static and maximum cyclic heel and forefoot reference loading .....</b>	<b>1</b>
2.2.1 <b>Basic relationships and conditions .....</b>	<b>1</b>
2.2.2 <b>Lines of action of the resultant reference forces FR1 and FR2 .....</b>	<b>2</b>
2.2.3 <b>Position of the top load application point PT .....</b>	<b>3</b>
2.3 <b>Magnitudes of static and maximum cyclic heel and forefoot reference loading .....</b>	<b>6</b>
2.4 <b>Reference test loading conditions of static and cyclic tests .....</b>	<b>7</b>
2.4.1 <b>Static tests .....</b>	<b>7</b>
2.4.2 <b>Cyclic test .....</b>	<b>8</b>
<b>3</b> <b>Guidance on the design of appropriate test equipment for the application of ISO 22675 ...</b>	<b>20</b>
3.1 <b>Background statement .....</b>	<b>20</b>
3.2 <b>Basic design for test equipment .....</b>	<b>20</b>
3.3 <b>Design variants for load application .....</b>	<b>24</b>
3.3.1 <b>General .....</b>	<b>24</b>
3.3.2 <b>Design variant A .....</b>	<b>24</b>
3.3.3 <b>Design variant B .....</b>	<b>24</b>
3.3.4 <b>Main differences between design variants A and B .....</b>	<b>25</b>
3.4 <b>Examples of crank gear designs .....</b>	<b>25</b>
3.4.1 <b>General .....</b>	<b>25</b>
3.4.2 <b>Asymmetrical (60:40) crank gear .....</b>	<b>25</b>
3.4.3 <b>Symmetrical (50:50) crank gear .....</b>	<b>26</b>
3.5 <b>Effect of deviations of the tilting angle (<math>t</math>) from the specified profile (curve), addressed in 3.4, on the test loading conditions of ISO 22675 .....</b>	<b>30</b>
3.6 <b>Effect of the position of the tilting axis TA of the foot platform on the elevation E and the A-P displacement <math>f</math> of the test sample at the foot .....</b>	<b>35</b>
3.6.1 <b>General .....</b>	<b>35</b>
3.6.2 <b>Position of the tilting axis TA of the foot platform .....</b>	<b>36</b>
3.6.3 <b>Values of elevation E .....</b>	<b>36</b>
3.6.4 <b>Values of A-P displacement <math>f</math> .....</b>	<b>37</b>
3.6.5 <b>Conclusions .....</b>	<b>39</b>
3.7 <b>Effect of the elevation E and A-P displacement <math>f</math> of the test sample, caused by the tilting of the foot platform, on the test loading conditions of ISO 22675 .....</b>	<b>43</b>
3.8 <b>Transposition of the top load application point PT for compensation of the dependence of the position of the tilting axis TA of the foot platform on the foot length L .....</b>	<b>49</b>
3.8.1 <b>General .....</b>	<b>49</b>
3.8.2 <b>Possibilities of transposing the top load application point PT .....</b>	<b>49</b>
3.8.3 <b>Practicality .....</b>	<b>50</b>
3.9 <b>Effect of the position of the tilting axis TA of the foot platform on the tilting moment and the driving torque .....</b>	<b>53</b>
3.10 <b>Alternative design of foot platform .....</b>	<b>58</b>

<b>Annex A (informative) Information on ISO 22675 .....</b>	<b>61</b>
<b>Bibliography .....</b>	<b>62</b>
<b>Figure 1 -- Illustration of different components of loading .....</b>	<b>10</b>
<b>Figure 2 -- Profiles (curves) of force components and tilting angle for test loading level P5, based on gait analysis data representative of normal level walking .....</b>	<b>11</b>
<b>Figure 3 -- Profiles (curves) of force components and angles for test loading level P5, establishing the basis from which to specify the test loading conditions of ISO 22675 .....</b>	<b>12</b>
<b>Figure 4 -- Illustration of different test loading conditions for test loading level P5 .....</b>	<b>13</b>
<b>Figure 5 -- Illustration of different test loading conditions for test loading levels P5, P4 and P3 .....</b>	<b>15</b>
<b>Figure 6 -- Illustration of the dependence of the position of the top load application point PT on the foot length L (see 2.2.3) .....</b>	<b>17</b>
<b>Figure 7 -- Illustration of the progression of the line of action of the resultant force FR from heel contact to toe-off in 30 ms time increments for related values of angle shown in Figure 3 .....</b>	<b>19</b>
<b>Figure 8 -- Diagrammatic view of test equipment with test sample .....</b>	<b>22</b>
<b>Figure 9 -- Parameters of a crank gear capable of driving the foot platform of the test equipment to generate the profile (curve) (t) .....</b>	<b>23</b>
<b>Figure 10 -- Asymmetrical (60:40) crank gear according to 3.4.2 -- Tilting range - 20° (heel contact) to + 40° (toe-off) .....</b>	<b>27</b>
<b>Figure 11 -- Symmetrical (50:50) crank gear according to 3.4.3 -- Tilting range -- 20° (heel contact) via + 40° (toe-off) to + 50° .....</b>	<b>28</b>
<b>Figure 12 -- Tilting characteristics of asymmetrical (60:40) crank gear according to 3.4.2 and Figure 10 and symmetrical (50:50) crank gear according to 3.4.3 and Figure 11 .....</b>	<b>29</b>
<b>Figure 13 -- Profiles (curves) of angles , and as specified and as produced by crank gear 60:40 ....</b>	<b>31</b>
<b>Figure 14 -- Illustration of angular deviations produced by crank gear 60:40 .....</b>	<b>32</b>
<b>Figure 15 -- Profiles (curves) of force components FP and FT, as specified and as produced by crank gear 60:40 .....</b>	<b>33</b>
<b>Figure 16 -- Illustration of force deviations produced by crank gear 60:40 .....</b>	<b>34</b>
<b>Figure 17 -- Illustration of distortion of time base of test force F produced by crank gear 60:40 .....</b>	<b>35</b>
<b>Figure 18 -- Effect of f-position of tilting axis TA of foot platform on the elevation E of the foot at the instants of heel contact and toe-off .....</b>	<b>40</b>
<b>Figure 19 -- Effect of u-position of tilting axis TA of foot platform on the A/P displacement f of the foot at the instant of toe-off .....</b>	<b>41</b>
<b>Figure 20 -- Values of elevation E and A-P displacement f at specific positions of tilting axis TA .....</b>	<b>42</b>
<b>Figure 21 -- Illustration of the effect of A-P displacement f on the angular movement of the test sample about the "internal" top load application point PT in an arrangement according to 3.3.2 .....</b>	<b>46</b>

Figure 22 -- Illustration of the effect of A-P displacement $f$ on the angular movement of the test sample about the "external" top load application point PTE in an arrangement according to 3.3.3 .....	47
Figure 23 -- Illustration of possibilities of transposing the top load application point PT for compensating the dependence of the position of the tilting axis TA of the foot platform on the foot length L .....	51
Figure 24 -- Illustration of the effect of a fixed compromise offset $u_{TA}$ , C of the tilting axis TA of the foot platform on the A-P displacement $f$ at the foot for different foot lengths L [see 3.8.2 c) 2)] .....	52
Figure 25 -- Illustration of effective lever arms .....	55
Figure 26 -- Force transmission by asymmetrical (60:40) crank gear drive according to 3.4.2 and Figure 10 .....	56
Figure 27 -- Force transmission by symmetrical (50:50) crank gear drive according to 3.4.3 and Figure 11 .....	57
Figure 28 -- Tilting characteristic of foot platform of polycentric (four-bar-linkage) design .....	59
Figure 29 -- Horizontal displacement of instantaneous centre IC of foot platform of polycentric (four-bar-linkage) design .....	60
Table 1 -- Magnitudes of resultant reference forces $FR_{1x}$ and $FR_{2x}$ .....	7
Table 2 -- Coordinates $f_{TA}$ and $u_{TA}$ of the tilting axis TA of the foot platform and related values of elevation E and A-P displacement $f$ for foot length $L = 30$ cm .....	43
Table 3 -- Specific values demonstrating the effect of A-P displacement $f$ on the angular movement of the test sample about the top load application point PT for foot length $L = 30$ cm .....	48
Table 4 -- Possibilities of transposing the top load application point PT for compensating the dependence of the position of the tilting axis TA of the foot platform on the foot length L .....	53
Table 5 -- Moments at tilting axis TA and crankshaft CS, generated by test force $F(t)$ at test loading level P5, applied to test sample of foot length $L = 30$ cm .....	58
Table A.1 -- Excerpt from contents of Annexes A and E of ISO 22675:2006 and list of corresponding clause/s of this Technical Report, in which selected items are dealt with ..	61