DIN EN ISO 11137-2:2007-09 (E)

Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose (ISO 11137-2:2006, corrected version 2006-08-01)

Contents

Page

Foreword 4

Introduction 5

1 Scope 6

2 Normative references 6

3 Abbreviations, terms and definitions 6
3.1 Abbreviations 6
3.2 Terms 8

4 Definition and maintenance of product families for dose setting, dose substantiation and sterilization dose auditing 9
4.1 General 9
4.2 Defining product families 9
4.3 Designation of product to represent a product family for performance of a verification dose experiment or sterilization dose audit 10
4.4 Maintaining product families 11
4.5 Effect of failure of establishment of sterilization dose or of a sterilization dose audit on a product family 12

5 Selection and testing of product for establishing and verifying the sterilization dose 12
5.1 Nature of product 12
5.2 Sample item portion (SIP) 13
5.3 Manner of sampling 13
5.4 Microbiological testing 14
5.5 Irradiation 14

6 Methods of dose establishment 14

7 Method 1: dose setting using bioburden information 15
7.1 Rationale 15
7.2 Procedure for Method 1 for product with an average bioburden 1,0 for multiple production batches 16
7.3 Procedure for Method 1 for product with an average bioburden 1,0 for a single production batch 21
7.4 Procedure for Method 1 for product with an average bioburden in the range 0,1 to 0,9 for multiple or single production batches 23

8 Method 2: Dose setting using fraction positive information from incremental dosing to determine an extrapolation factor 23
8.1 Rationale 23
8.2 Procedure for Method 2A 24
8.3 Procedure for Method 2B 27

9 Method VDmax -- Substantiation of 25 kGy or 15 kGy as the sterilization dose 30
9.1 Rationale 30
9.2 Procedure for Method VDmax 25 for multiple production batches 31
9.3 Procedure for Method VDmax 25 for a single production batch 34
9.4 Procedure for Method VDmax 15 for multiple production batches 35

9.5 Procedure for Method VDmax 15 for a single production batch 38

10 Auditing sterilization dose 39
10.1 Purpose and frequency 39
10.2 Procedure for auditing a sterilization dose established using Method 1 or Method 2 40
10.3 Procedure for auditing a sterilization dose substantiated using VDmax 42

11 Worked examples 46
11.1 Worked examples for Method 1 46
11.2 Worked examples for Method 2 49
11.3 Worked examples for Method VDmax 58
11.4 Worked example of a sterilization dose audit for a dose established using Method 1, the findings from which necessitated augmentation of the sterilization dose 60
11.5 Worked example of a sterilization dose audit for a dose established using Method 2A, the findings from which necessitated augmentation of the sterilization dose 61
11.6 Worked example of a sterilization dose audit for a sterilization dose substantiated using Method VDmax 25 63

Bibliography 64

Requirements of EU Directives 90/385/EEC of 20 June 1990 concerning active implantable medical devices, 93/42/EEC of 14 June 1993 concerning medical devices and 98/79/EC of 7 December 1988 concerning in vitro diagnostic medical devices 66