

ISO 23747:2007-07 (E)

Anaesthetic and respiratory equipment - Peak expiratory flow meters for the assessment of pulmonary function in spontaneously breathing humans

| Contents | | Page |
|--|--|-------------|
| Foreword | | iv |
| Introduction | | v |
| 1 | Scope | 1 |
| 2 | Normative references | 1 |
| 3 | Terms and definitions | 1 |
| 4 | General requirements | 2 |
| 4.1 | Safety for PEFMs that utilize electricity | 2 |
| 4.2 | Mechanical safety for all PEFMs | 2 |
| 5 | Identification, marking and documents | 2 |
| 5.1 | Marking of the scale or display | 2 |
| 5.2 | Marking of PEFM or packaging | 3 |
| 5.3 | Instructions for use | 3 |
| 5.4 | Technical description | 4 |
| 6 | PEFM measurement range | 4 |
| 7 | Performance requirements | 4 |
| 7.1 | Error of measurement | 4 |
| 7.2 | Linearity | 4 |
| 7.3 | Resistance to flow | 5 |
| 7.4 | Frequency response | 5 |
| 8 | Dismantling and reassembly | 5 |
| 9 | Effects of mechanical ageing | 5 |
| 10 | Effects of dropping a hand-held PEFM | 5 |
| 11 | Cleaning, sterilization and disinfection | 5 |
| 11.1 | Re-usable PEFM and parts | 5 |
| 11.2 | PEFM and parts delivered sterile | 6 |
| 12 | Compatibility with substances | 6 |
| 13 | Biocompatibility | 6 |
| Annex A (informative) Rationale for tests and examples of test apparatus | | 7 |
| Annex B (normative) Determination of error, repeatability and resistance to PEFM output | | 10 |
| Annex C (normative) Determination of frequency response | | 13 |
| Annex D (normative) Test methods for determination of the effects of dismantling, ageing and dropping | | 15 |
| Annex E (informative) Environmental aspects | | 17 |
| Annex F (informative) Reference to the essential principals | | 19 |
| Bibliography | | 22 |