ISO 15882:2008-09 (E)

Sterilization of health care products - Chemical indicators - Guidance for selection, use and interpretation of resul ts

Contents

Page

Foreword iv

Introduction v

1 Scope 1

2 Terms and definitions 1

3 General considerations 3

4 Classes of chemical indicator 5
4.1 General 5
4.2 Class 1: Process indicators 5
4.3 Class 2: Indicators for use in specific tests 6
4.4 Class 3: Single variable indicators 6
4.5 Class 4: Multi-variable indicators 8
4.6 Class 5: Integrating indicators 8
4.7 Class 6: Emulating indicators 9

5 Selection of chemical indicators 10

6 Use of chemical indicators 10
6.1 Class 1 process indicators 10
6.2 Class 2 indicators 11
6.3 Class 3, 4, 5 and 6 indicators 11
6.4 Indicators for use with process challenge devices 11

7 Interpretation of results from chemical indicators 12
7.1 General 12
7.2 Chemical indicator responses 12
7.3 Chemical indicators showing "fail" response 12

8 Chemical indicators in sterility assurance procedures 12
8.1 General 12
8.2 Record keeping 13

9 Personnel training 13

10 Storage and handling 14
11 Labelling 14
11.1 General 14
11.2 Indicator marking 14
11.3 Process marking 14
11.4 Package marking 14

Annex A (informative) Background on the Bowie and Dick test 16

Annex B (informative) Explanation of the terms "parameter" and "variable" 19

Annex C (informative) Rationale for the requirements for integrating indicators and the link to the requirements for biological indicators (BIs) specified in the ISO 11138 series and microbial inactivation (derived from ISO 11140-1) 20

Annex D (informative) Specifications for porosity 27

Annex E (informative) Figure showing relationship of indicator components 29

Bibliography 30