ISO 80601-2-61:2011-04 (E)

Medical electrical equipment - Part 2-61: Particular requirements for b asic safety and essential performance of pulse oximeter equipment

Contents

Page

Foreword vi

Introduction viii

201.1 Scope, object and related standards 1
201.1. 1 * Scope 1
201.1. 2 Object 1
201.1. 3 Collateral standards 2
201.1. 4 Particular standards 2

201.2 Normative references 3

201.3 Terms and definitions 4

201.4 General requirements 8
201.4. 3 ESSENTIAL PERFORMANCE 8
201.4. 101 * Additional requirements for ESSENTIAL PERFORMANCE 8
201.4. 102 Additional requirements for acceptance criteria 9
201.4. 103 Additional requirements for PULSE OXIMETER EQUIPMENT, parts and ACCESSORIES 9

201.5 General requirements for testing of ME EQUIPMENT 9

201.6 Classification of ME EQUIPMENT and ME SYSTEMS 10

201.7 ME EQUIPMENT identification, marking and documents 10
201.7. 2.3 Consult ACCOMPANYING DOCUMENTS 10
201.7. 2.9 IP classification 10
201.7. 2.101 Additional requirements for marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts 10
201.7. 2.4.101 Additional requirements for ACCESSORIES 11
201.7. 2.13.101 Additional requirements for physiological effects 11
201.7. 2.17.101 Additional requirements for protective packaging 11
201.7. 4.3 Unit of measure 12
201.7. 9.1 Additional general requirements 12
201.7. 9.2.1.101 Additional general requirements 12
201.7. 9.2.2.101 Additional requirements for warnings and safety notices 13
201.7. 9.2.8.101 Additional requirements for start-up PROCEDURE 13
201.7. 9.2.9.101 Additional requirements for operating instructions 13
201.7. 9.2.14.101 Additional requirements for ACCESSORIES, supplementary equipment, used material 14
201.7. 9.3.1.101 * Additional general requirements 14

201.8 Protection against electrical HAZARDS from ME EQUIPMENT 15
201.8. 3.101 Additional requirements for classification of APPLIED PARTS 15

201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS 15

201.10 Protection against unwanted and excessive radiation HAZARDS 15

201.11 Protection against excessive temperatures and other HAZARDS 15
201.11. 6.5.101 * Additional requirements for ingress of water or particulate matter into ME EQUIPMENT or ME SYSTEM 16
201.11. 8.101 Additional requirements for interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT 16
201.11. 8.101.1 Supply failure TECHNICAL ALARM CONDITION 16
201.11. 8.101.2 Settings and data storage following short interruptions or automatic switchover 16
201.11. 8.101.3 Operation following long interruptions 17

201.12 ACCURACY of controls and instruments and protection against hazardous outputs 17

201.12. 1 ACCURACY of controls and instruments 17
201.12. 1.101 * SpO_2 accuracy of pulse oximeter equipment 17
201.12. 1.101.1 * Specification 17
201.12. 1.101.2 Determination of SpO_2 ACCURACY 18
201.12. 1.101.2.1 * Data collection 18
201.12. 1.101.2.2 * Data analysis 18
201.12. 1.101.2.3 Characteristics of the clinical study population 19
201.12. 1.102 Accuracy under conditions of motion 19
201.12. 1.103 ACCURACY under conditions of low perfusion 19
201.12. 1.104 Pulse rate ACCURACY 19
201.12. 4 Protection against hazardous output 19
201.12. 4.101 * DATA UPDATE PERIOD 19
201.12. 4.102 * Signal inadequacy 20
201.13 HAZARDOUS SITUATIONS and fault conditions 20
201.13. 101 Detection of PULSE OXIMETER PROBE faults and PROBE CABLE EXTENDER faults 20
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) 20
201.15 Construction of ME EQUIPMENT 21
201.15. 3.5.101 * Additional requirements for rough handling 21
201.15. 3.5.101.1 * Shock and vibration 21
201.15. 3.5.101.2 * Shock and vibration for professional transport 22
201.15. 101 Mode of operation 23
201.16 ME SYSTEMS 23
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS 23
201.101 * PULSE OXIMETER PROBES and PROBE CABLE EXTENDERS 23
201.101. 1 General 23
201.101. 2 Labelling 24
201.102 Saturation pulse INFORMATION SIGNAL 24
201.103 SIGNAL INPUT/OUTPUT PART 24
201.103. 1 General 24
201.103. 2 Connection to electronic health record 24
201.103. 3 Connection to a distributed alarm system 25
201.103. 4 Connection for remote control 25
202 Medical electrical equipment — Part 1-2: General requirements for safety — Collateral standard: Electromagnetic compatibility — Requirements and tests 25
202.6.2.1.1 IMMUNITY TEST LEVELS 25
202.6.2.1.7 PATIENT simulation 25
202.6.2.1.10 * Requirements 26
202.6.2.3 * Radiated RF electromagnetic fields 26
208 Medical electrical equipment — Part 1-8: General requirements for safety — Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems 26
208.6.1.2.101 * Additional requirements for ALARM CONDITION priority 26
208.6.5.4.101 * Additional requirements for DEFAULT ALARM PRESET 27
208.6.8.5.101 Additional requirements for ALARM SIGNAL inactivation states, indication and access 27
Annex C (informative) **Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS** 28
Annex D (informative) **Symbols on marking** 32
Annex AA (informative) **Particular guidance and rationale** 34

Annex BB (informative) **Skin temperature at the PULSE OXIMETER PROBE** **41**

Annex CC (informative) **Determination of ACCURACY** **45**

Annex DD (informative) **Calibration standards** **53**

Annex EE (informative) **Guideline for evaluating and documenting SpO_2 ACCURACY in human subjects** **54**

Annex FF (informative) **Simulators, calibrators and FUNCTIONAL TESTERS for PULSE OXIMETER EQUIPMENT** **61**

Annex GG (informative) **Concepts of ME EQUIPMENT response time** **70**

Annex HH (informative) **Reference to the essential principles of safety and performance of medical devices in accordance with ISO/TR 16142** **74**

Bibliography **76**

Alphabetized index of defined terms used in this standard **81**