

ISO 1135-4:2012-03 (E)

Transfusion equipment for medical use - Part 4: Transfusion sets for single use

| Contents | | Page |
|---|--|-------------|
| Foreword | | iv |
| 1 | Scope | 1 |
| 2 | Normative references | 1 |
| 3 | General requirements | 2 |
| 3.1 | Nomenclature for components of the transfusion set | 2 |
| 3.2 | Maintenance of sterility | 2 |
| 3.3 | Designation | 3 |
| 4 | Materials | 3 |
| 5 | Physical requirements | 3 |
| 5.1 | Particulate contamination | 3 |
| 5.2 | Leakage | 3 |
| 5.3 | Tensile strength | 3 |
| 5.4 | Closure-piercing device | 3 |
| 5.5 | Tubing | 4 |
| 5.6 | Filter for blood and blood components | 4 |
| 5.7 | Drip chamber and drip tube | 4 |
| 5.8 | Flow regulator | 4 |
| 5.9 | Flow rate of blood and blood components | 4 |
| 5.10 | Injection site | 5 |
| 5.11 | Male conical fitting | 5 |
| 5.12 | Protective caps | 5 |
| 6 | Chemical requirements | 5 |
| 6.1 | Reducing (oxidizable) matter | 5 |
| 6.2 | Metal ions | 5 |
| 6.3 | Titration acidity or alkalinity | 5 |
| 6.4 | Residue on evaporation | 5 |
| 6.5 | UV absorption of extract solution | 5 |
| 7 | Biological requirements | 6 |
| 7.1 | General | 6 |
| 7.2 | Sterility | 6 |
| 7.3 | Pyrogenicity | 6 |
| 7.4 | Haemolysis | 6 |
| 7.5 | Toxicity | 6 |
| 8 | Labelling | 6 |
| 8.1 | Unit container | 6 |
| 8.2 | Shelf or multi-unit container | 7 |
| 9 | Packaging | 7 |
| 10 | Disposal | 7 |
| Annex A (normative) Physical tests | | 8 |
| Annex B (normative) Chemical tests | | 12 |

| | |
|---|-----------|
| Annex C (normative) Biological tests | 14 |
| Bibliography | 15 |