

# DIN EN ISO 18113-5:2013-01 (E)

## In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 5: In vitro diagnostic instruments for self-testing (ISO 18113-5:2009)

---

<b>Contents</b>		<b>Page</b>
Foreword .....		3
Introduction .....		4
1	Scope .....	5
2	Normative references .....	5
3	Terms and definitions .....	6
4	Essential requirements .....	6
5	Labels and marking .....	6
5.1	General .....	6
5.2	Identification of the IVD instrument .....	6
6	Elements of the instructions for use .....	7
7	Content of the instructions for use .....	7
7.1	Manufacturer .....	7
7.2	Identification of the IVD instrument .....	7
7.3	Intended use .....	8
7.4	Storage and handling .....	8
7.5	Warnings and precautions .....	8
7.6	Instrument installation .....	8
7.7	Principles of measurement .....	9
7.8	Performance of the IVD instrument .....	9
7.9	Limitations of use .....	9
7.10	Preparation prior to operation .....	9
7.11	Operating procedure .....	9
7.12	Control procedure .....	10
7.13	Reading of examination results .....	10
7.14	Special functions .....	10
7.15	Shut-down procedure .....	10
7.16	Disposal information .....	10
7.17	Maintenance .....	11
7.18	Troubleshooting .....	11
7.19	Follow-up action .....	11
Bibliography .....		12
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of the EU Directive 98/79/EC on "in vitro Diagnostic Medical Devices" .....		13