

DIN EN ISO 10555-1:2013-11 (E nglisch)

Intravascular catheters - Sterile and single-use catheters - Part 1: General requirements (ISO 10555-1:2013)

| Contents | | Page |
|-------------------------------|--|-------------|
| Foreword | | 3 |
| 1 | Scope | 4 |
| 2 | Normative references | 4 |
| 3 | Terms and definitions | 4 |
| 4 | Requirements | 7 |
| 4.1 | General | 7 |
| 4.2 | Radio-detectability | 7 |
| 4.3 | Biocompatibility | 7 |
| 4.4 | Surface | 7 |
| 4.5 | Corrosion resistance | 7 |
| 4.6 | Peak tensile force | 7 |
| 4.7 | Freedom from leakage | 8 |
| 4.8 | Hubs | 8 |
| 4.9 | Flowrate | 8 |
| 4.10 | Power injection | 8 |
| 4.11 | Side holes | 8 |
| 4.12 | Distal tip | 8 |
| 5 | Designation of nominal size | 8 |
| 5.1 | Outside diameter | 8 |
| 5.2 | Nominal effective length | 9 |
| 6 | Information to be supplied by the manufacturer | 9 |
| 6.1 | General | 9 |
| 6.2 | Marking on the device and/or primary packaging | 9 |
| 6.3 | Instructions for use | 10 |
| 6.4 | Marking on the secondary packaging | 10 |
| Annex A (normative) | Test method for corrosion resistance | 11 |
| Annex B (normative) | Method for determining peak tensile force | 12 |
| Annex C (normative) | Test method for liquid leakage under pressure | 14 |
| Annex D (normative) | Test method for air leakage into hub assembly during aspiration | 16 |
| Annex E (normative) | Determination of flowrate through catheter | 18 |
| Annex F (normative) | Test for burst pressure under static conditions | 20 |
| Annex G (normative) | Power injection test for flowrate and device pressure (only for products indicated for power injection) | 22 |
| Annex H (informative) | Units of measurement systems other than those specified in this part of ISO 10555, which may additionally be used | 25 |
| Annex ZA (informative) | Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC amended by Directive 2007/47/EEC | 27 |
| Bibliography | | 29 |