ISO/IEEE 11073-10421:2012-11 (E)

Health informatics - Personal health device communication - Part 10421: Devic e specialization - Peak expiratory flow monitor (peak flow)

Contents

Page

1. Overview 1
1.1 Scope 1
1.2 Purpose 2
1.3 Context 2

2. Normative references 2

3. Definitions, acronyms, and abbreviations 2
3.1 Definitions 3
3.2 Acronyms and abbreviations 3

4.1 General 4

5. Peak expiratory flow monitor device concepts and modalities 5
5.1 General 5
5.2 PEF 5

6. Peak expiratory flow monitor domain information model 7
6.1 Overview 7
6.2 Class extensions 7
6.3 Object instance diagram 7
6.4 Types of configuration 9
6.5 Medical device system object 9
6.6 Numeric objects 13
6.7 Real-time sample array objects 20
6.8 Enumeration objects 20
6.9 PM-store objects 22
6.10 Scanner objects 22
6.11 Class extension objects 22
6.12 Peak expiratory flow monitor information model extensibility rules 22

7. Peak expiratory flow monitor service model 22
7.1 General 22
7.2 Object access services 22
7.3 Object access event report services 23

8. Peak expiratory flow monitor communication model 24
8.1 Overview 24
8.2 Communications characteristics 24
8.3 Association procedure 24
8.4 Configuring procedure 26
8.5 Operating procedure 28
8.6 Time synchronization 29

9. Test associations 29
9.1 Behavior with standard configuration 29
9.2 Behavior with extended configurations 30

10. Conformance 30
10.1 Applicability 30
10.2 Conformance specification 30

10.3 Levels of conformance 30
10.4 Implementation conformance statements 31

Annex A (informative) Bibliography 36

Annex B (normative) Any additional ASN.1 definitions 37

Annex C (normative) Allocation of identifiers 38

Annex D (informative) Message sequence examples 39

Annex E (informative) Protocol data unit examples 41

Annex F (informative) IEEE list of participants 52