DIN EN ISO 11137-1:2013-12 (E)

Ster ilization of health care products - Radiation - Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices (ISO 11137-1:2006 + Amd 1:2013) (includes Amendment A1:2013)

Contents

Page

Introduction 5

1 Scope 7

2 Normative references 8

3 Terms and definitions 8

4 Quality management system elements 14
4.1 Documentation 14
4.2 Management responsibility 15
4.3 Product realization 15
4.4 Measurement, analysis and improvement -- Control of nonconforming product 15

5 Sterilizing agent characterization 15
5.1 Sterilizing agent 15
5.2 Microbicidal effectiveness 15
5.3 Material effects 15
5.4 Environmental considerations 16

6 Process and equipment characterization 16
6.1 Process 16
6.2 Equipment 16

7 Product definition 17

8 Process definition 18
8.1 Establishing the maximum acceptable dose 18
8.2 Establishing the sterilization dose 18
8.3 Specifying the maximum acceptable dose and the sterilization dose 19
8.4 Transference of maximum acceptable, verification or sterilization dose between radiation sources 19

9 Validation 20
9.1 Installation qualification 20
9.2 Operational qualification 20
9.3 Performance qualification 21
9.4 Review and approval of validation 21

10 Routine monitoring and control 22

11 Product release from sterilization 23

12 Maintaining process effectiveness 23
12.1 Demonstration of continued effectiveness 23
12.2 Recalibration 26
12.3 Maintenance of equipment 26
12.4 Requalification of equipment 26
12.5 Assessment of change 26

Annex A (informative) Guidance 27

Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 90/385/EEC on Active Implantable Medical Devices 42

Annex ZB (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on Medical Devices 43

Annex ZC (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 98/79/EC on in vitro diagnostic medical devices 44

Bibliography 45