ISO/TS 13004:2013-05 (E)

Sterilization of health care products - Radiation - Substantiation of selected sterilization dose: Method VD<(Inde x)max><(hoch)SD>

Contents

Page

Foreword iv

Introduction v

1 Scope 1
1.1 Inclusions 1
1.2 Exclusions 1
1.3 Application 1

2 Normative references 1

3 Termsanddefinitions 1

4 Definitionandmaintenanceofproductfamiliesforsterilizationdosesubstantiationand sterilization dose auditing 5
4.1 General 5
4.2 Defining product families 5
4.3 Designation of product to represent a product family 6
4.4 Maintaining product families 7
4.5 Consequence of failure of sterilization dose substantiation or sterilization dose audit 8

5 Selection and testing of product for substantiating and auditing a selected sterilization dose 8
5.1 Nature of product 8
5.2 Sample item portion (SIP) 9
5.3 Manner of sampling 10
5.4 Microbiological testing 11
5.5 Irradiation 11

6 Method VD SD max -- Substantiation of a selected sterilization dose of 17,5, 20, 22,5, 27,5, 30, 32,5, or 35 kGy 11
6.1 Rationale 11
6.2 Procedure for Method VD SD max for multiple production batches 12
6.3 Procedure for Method VD SD max for a single production batch 17

7 Maintaining process effectiveness 21
7.1 General 21
7.2 Determination of bioburden 22
7.3 Sterilization dose audit 22

8 Tables of values for SIP equal to 1,0 VDmaxSD, SIP dose reduction factor and augmentation dose corresponding to applicable values of average bioburden for selected sterilization doses of 17,5, 20, 22,5, 27,5, 30, 32,5 and 35 kGy 27

9 Worked examples 53
9.1 Substantiation of a selected sterilization dose of 17,5 kGy (SIP less than 1,0) 53
9.2 Substantiation of a selected sterilization dose of 30 kGy (SIP equal to 1,0) 54
9.3 Sterilization dose audit for a sterilization dose substantiated using Method VDmax22,5, the findings from which necessitated augmentation of the sterilization dose 55

Bibliography 57