

ISO/TS 13004:2013-05 (E)

Sterilization of health care products - Radiation - Substantiation of selected sterilization dose: Method VD<(Inde x)max><(hoch)SD>

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
1.1 Inclusions	1
1.2 Exclusions	1
1.3 Application	1
2 Normative references	1
3 Terms and definitions	1
4 Definition and maintenance of product families for sterilization dose substantiation and sterilization dose auditing	5
4.1 General	5
4.2 Defining product families	5
4.3 Designation of product to represent a product family	6
4.4 Maintaining product families	7
4.5 Consequence of failure of sterilization dose substantiation or sterilization dose audit	8
5 Selection and testing of product for substantiating and auditing a selected sterilization dose	8
5.1 Nature of product	8
5.2 Sample item portion (SIP)	9
5.3 Manner of sampling	10
5.4 Microbiological testing	11
5.5 Irradiation	11
6 Method VD SD max -- Substantiation of a selected sterilization dose of 17,5, 20, 22,5, 27,5, 30, 32,5, or 35 kGy	11
6.1 Rationale	11
6.2 Procedure for Method VD SD max for multiple production batches	12
6.3 Procedure for Method VD SD max for a single production batch	17
7 Maintaining process effectiveness	21
7.1 General	21
7.2 Determination of bioburden	22
7.3 Sterilization dose audit	22
8 Tables of values for SIP equal to 1,0 VDmaxSD, SIP dose reduction factor and augmentation dose corresponding to applicable values of average bioburden for selected sterilization doses of 17,5, 20, 22,5, 27,5, 30, 32,5 and 35 kGy	27
9 Worked examples	53
9.1 Substantiation of a selected sterilization dose of 17,5 kGy (SIP less than 1,0)	53
9.2 Substantiation of a selected sterilization dose of 30 kGy (SIP equal to 1,0)	54
9.3 Sterilization dose audit for a sterilization dose substantiated using Method VDmax22,5, the findings from which necessitated augmentation of the sterilization dose	55
Bibliography	57