ISO 11137-2:2013-06 (E)

Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose

Contents

Page

Foreword v

Introduction vi

1 Scope 1

2 Normative references 1

3 Terms,definitions,andabbreviatedterms 1
3.1 Terms and definitions 1
3.2 Abbreviated terms 3

4 Definitionandmaintenanceofproductfamiliesfordosesetting,dosesubstantiation,and sterilization dose auditing 4
4.1 General 4
4.2 Defining product families 4
4.3 Designation of product to represent a product family for performance of a verification dose experiment or sterilization dose audit 5
4.4 Maintaining product families 6
4.5 Effect of failure of establishment of sterilization dose or of a sterilization dose audit on a product family 7

5 Selection and testing of product for establishing the sterilization dose 7
5.1 Nature of product 7
5.2 Sample item portion (SIP) 8
5.3 Manner of sampling 9
5.4 Microbiological testing 9
5.5 Irradiation 9

6 Methods of dose establishment 9

7 Method 1: dose setting using bioburden information 10
7.1 Rationale 10
7.2 Procedure for Method 1 for product with an average bioburden greater than or equal to 1,0 for multiple production batches 11
7.3 Procedure for Method 1 for product with an average bioburden greater than or equal to 1,0 for a single production batch 17
7.4 Procedure for Method 1 for product with an average bioburden in the range 0,1 to 0,9 for multiple or single production batches 19

8 Method 2: Dose setting using fraction positive information from incremental dosing to determine an extrapolation factor 20
8.1 Rationale 20
8.2 Procedure for Method 2A 21
8.3 Procedure for Method 2B 24

9 Method VDmax -- Substantiation of 25 kGy or 15 kGy as the sterilization dose 28
9.1 Rationale 28
9.2 Procedure for Method VDmax25 for multiple production batches 29
9.3 Procedure for Method VDmax25 for a single production batch 34
9.4 Procedure for Method VDmax15 for multiple production batches 37

9.5 Procedure for Method VDmax15 for a single production batch 40

10 Sterilization dose audit 43
10.1 Purpose and frequency 43
10.2 Procedure for auditing a sterilization dose established using Method 1, Method 2A, or Method 2B 43
10.3 Procedure for auditing a sterilization dose substantiated using Method VDmax25 or Method VDmax15 46
10.4 Failure of a sterilization dose audit 52

11 Worked examples 52
11.1 Worked examples for Method 1 52
11.2 Worked examples for Method 2 54
11.3 Worked examples for Method VDmax 62
11.4 Worked example of a sterilization dose audit for a dose established using Method 1, the findings from which necessitated augmentation of the sterilization dose 64
11.5 Worked example of a sterilization dose audit for a dose established using Method 2A, the findings from which necessitated augmentation of the sterilization dose 64
11.6 Worked example of a sterilization dose audit for a sterilization dose substantiated using Method VDmax25 65

Bibliography 67