

ISO 11137-2:2013-06 (E)

Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose

Contents		Page
Foreword		v
Introduction		vi
1	Scope	1
2	Normative references	1
3	Terms, definitions, and abbreviated terms	1
3.1	Terms and definitions	1
3.2	Abbreviated terms	3
4	Definition and maintenance of product families for dose setting, dose substantiation, and sterilization dose auditing	4
4.1	General	4
4.2	Defining product families	4
4.3	Designation of product to represent a product family for performance of a verification dose experiment or sterilization dose audit	5
4.4	Maintaining product families	6
4.5	Effect of failure of establishment of sterilization dose or of a sterilization dose audit on a product family	7
5	Selection and testing of product for establishing the sterilization dose	7
5.1	Nature of product	7
5.2	Sample item portion (SIP)	8
5.3	Manner of sampling	9
5.4	Microbiological testing	9
5.5	Irradiation	9
6	Methods of dose establishment	9
7	Method 1: dose setting using bioburden information	10
7.1	Rationale	10
7.2	Procedure for Method 1 for product with an average bioburden greater than or equal to 1,0 for multiple production batches	11
7.3	Procedure for Method 1 for product with an average bioburden greater than or equal to 1,0 for a single production batch	17
7.4	Procedure for Method 1 for product with an average bioburden in the range 0,1 to 0,9 for multiple or single production batches	19
8	Method 2: Dose setting using fraction positive information from incremental dosing to determine an extrapolation factor	20
8.1	Rationale	20
8.2	Procedure for Method 2A	21
8.3	Procedure for Method 2B	24
9	Method VD_{max} -- Substantiation of 25 kGy or 15 kGy as the sterilization dose	28
9.1	Rationale	28
9.2	Procedure for Method VD_{max25} for multiple production batches	29
9.3	Procedure for Method VD_{max25} for a single production batch	34
9.4	Procedure for Method VD_{max15} for multiple production batches	37

9.5	Procedure for Method VDmax15 for a single production batch	40
10	Sterilization dose audit	43
10.1	Purpose and frequency	43
10.2	Procedure for auditing a sterilization dose established using Method 1, Method 2A, or Method 2B	43
10.3	Procedure for auditing a sterilization dose substantiated using Method VDmax25 or Method VDmax15	46
10.4	Failure of a sterilization dose audit	52
11	Worked examples	52
11.1	Worked examples for Method 1	52
11.2	Worked examples for Method 2	54
11.3	Worked examples for Method VDmax	62
11.4	Worked example of a sterilization dose audit for a dose established using Method 1, the findings from which necessitated augmentation of the sterilization dose	64
11.5	Worked example of a sterilization dose audit for a dose established using Method 2A, the findings from which necessitated augmentation of the sterilization dose	64
11.6	Worked example of a sterilization dose audit for a sterilization dose substantiated using Method VDmax25	65
	Bibliography	67