DIN EN 1422:2014-08 (E)

Sterilizers for medical purposes - Ethylene oxide sterilizers - Requirements and test methods

Contents

Page

Foreword 4

Introduction 5

1 Scope 6

2 Normative references 6

3 Terms and definitions 7

4 Technical requirements 12
4.1 General 12
4.1.1 Risk control and usability 12
4.1.2 Materials 13
4.2 Sterilizer chamber 13
4.2.1 Chamber size 13
4.2.2 Doors, closures and interlocks of the sterilizer chamber 13
4.2.3 Test connections 14
4.3 Design and construction 15
4.3.1 General 15
4.3.2 EO vaporizers 15
4.3.3 Pipework and fittings 15
4.3.4 Evacuation system 15
4.3.5 Control valves 16
4.3.6 Thermal insulation 16
4.3.7 Electrical and mechanical safety 16
4.3.8 Air or inert gas filter 16
4.3.9 Emission control 16
4.3.10 Framework and panelling 17
4.3.11 Loading equipment 17
4.3.12 Transport 17
4.4 Indicating, measuring, and recording instruments 17
4.4.1 General 17
4.4.2 Temperature sensor 18
4.4.3 Temperature indicating instruments 18
4.4.4 Pressure sensors 19
4.4.5 Timers and time indicating instruments 19
4.4.6 Sterilizing cycle counter 19
4.4.7 Relative humidity (RH) sensors 19
4.4.8 Ethylene Oxide (EO) concentration-measurement 19
4.4.9 Recording instruments 20
4.4.10 Indicating instruments 21

5 Process control 22
5.1 General 22
5.2 Software verification and validation 23
5.3 Sterilization cycle and automatic control 23
5.3.1 Automatic control 23
5.3.2 Sterilization cycle 24
5.4 Override of automatic control 27
5.5 Fault 27

6 Performance requirements 28
6.1 Sterilizing performance 28
6.1.1 Loading configuration 28
6.1.2 Physical parameters 28
6.1.3 Microbiological efficacy 28
6.2 EO removal (flushing) 29
6.3 Aeration 29

7 Sound power 29

8 Packaging, marking and labelling 29

9 Information to be supplied by the manufacturer 30

10 Service and local environment 32
10.1 General 32
10.2 Electricity 33
10.3 Sterilant 33
10.4 Circulation systems 33
10.5 Steam 33
10.6 Water 34
10.7 Air and inert gasses 34
10.8 Drainage and discharges 34
10.9 Ventilation and environment 34
10.10 Lighting 34

Annex A (normative) Test instrumentation 35

Annex B (normative) Leak test cycle 36

Annex C (normative) Sterilizer chamber profile testing 37

C.1 Sterilizer chamber internal surfaces 37
C.2 Empty sterilizer chamber 37

Annex D (normative) Microbiological test for EO sterilizers 38

D.1 General 38
D.2 Test equipment 38
D.3 Procedure 39
D.4 Interpretation of results 40

Annex E (informative) Environmental aspects 41

E.1 Environmental aspects regarding the life cycle of EO sterilizers 41
E.2 EO (brief description) 41
E.3 Environmental impact 41

Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices 44

Bibliography 48