

DIN EN ISO 11070:2015-03 (E)

Sterile single-use intravascular introducers, dilators and guidewires (ISO 11070:2014)

| Contents | | Page |
|--------------------|---|-------------|
| Foreword | | 4 |
| Introduction | | 5 |
| 1 | Scope | 6 |
| 2 | Normative references | 6 |
| 3 | Terms and definitions | 6 |
| 4 | General requirements | 10 |
| 4.1 | Sterilization | 10 |
| 4.2 | Biocompatibility | 10 |
| 4.3 | Surface | 10 |
| 4.4 | Corrosion resistance | 10 |
| 4.5 | Radio-detectability | 10 |
| 4.6 | Information to be supplied by the manufacturer | 10 |
| 5 | Additional requirements for introducer needles | 11 |
| 5.1 | General | 11 |
| 5.2 | Size designation | 11 |
| 5.3 | Needle point | 11 |
| 5.4 | Hub | 11 |
| 5.5 | Information to be supplied by the manufacturer | 11 |
| 6 | Additional requirements for introducer catheters | 11 |
| 6.1 | General | 11 |
| 6.2 | Tip | 12 |
| 6.3 | Peak tensile force | 12 |
| 6.4 | Hub | 12 |
| 6.5 | Size designation | 12 |
| 6.6 | Information to be supplied by the manufacturer | 13 |
| 7 | Additional requirements for sheath introducers | 13 |
| 7.1 | General | 13 |
| 7.2 | Size designation | 13 |
| 7.3 | Freedom from leakage from sheath introducer | 13 |
| 7.4 | Freedom from leakage through haemostasis valve | 13 |
| 7.5 | Hub | 13 |
| 7.6 | Peak tensile force | 13 |
| 7.7 | Information to be supplied by the manufacturer | 13 |
| 8 | Additional requirements for guidewires | 13 |
| 8.1 | General | 13 |
| 8.2 | Size designation | 14 |
| 8.3 | Safety wire | 14 |
| 8.4 | Fracture test | 14 |
| 8.5 | Flexing test | 14 |
| 8.6 | Peak tensile force of guidewire | 14 |
| 8.7 | Information to be supplied by the manufacturer | 15 |
| 9 | Additional requirements for dilators | 15 |

| | | |
|---|--|----|
| 9.1 | General | 15 |
| 9.2 | Size designation | 15 |
| 9.3 | Hub | 15 |
| 9.4 | Information to be supplied by the manufacturer | 15 |
| 10 | Additional requirements for kits containing combinations of devices specified in this International Standard | 15 |
| Annex A (informative) Guidance on materials and design | | 17 |
| Annex B (normative) Test method for corrosion resistance | | 18 |
| Annex C (normative) Method for determining peak tensile force of introducer catheters, sheath introducers, and dilators | | 19 |
| Annex D (normative) Test method for liquid leakage from sheath introducers under pressure | | 21 |
| Annex E (normative) Test method for liquid leakage through haemostasis valves of sheath introducers | | 22 |
| Annex F (normative) Test method for fracture of guidewires | | 23 |
| Annex G (normative) Test method for resistance of guidewires to damage by flexing | | 26 |
| Annex H (normative) Method for determining peak tensile force of guidewires | | 28 |
| Annex I (normative) Determination of strength of union of needle hub and needle | | 30 |
| Bibliography | | 31 |