ISO/TR 13570-2:2014-06 (E)

Wheelchairs - Part 2: Typical values and recommended limits of dimensions, mass and manoeuvring space as determined in ISO 7176-5

Contents

Page

Foreword iv

Introduction v

1 Scope 1

2 Normative references 1

3 Termsanddefinitions 1

4 Wheelchair groups 1
4.1 General 1
4.2 Wheelchairs with handrims 2
4.3 Electrically powered wheelchairs of class A 2
4.4 Electrically powered wheelchairs of class B 2
4.5 Electrically powered wheelchairs of class C 2
4.6 Electrically powered wheelchairs (scooter design) 2

5 Typical values and recommended limits for required measurements 2
5.1 General 2
5.2 Full overall length 2
5.3 Overall width 3
5.4 Handgrip height 4
5.5 Stowage length 5
5.6 Stowage width 6
5.7 Stowage height 7
5.8 Rising 8
5.9 Total mass 8
5.10 Mass of heaviest part 9
5.11 Pivot width 10
5.12 Reversing width 11
5.13 Turning diameter 12
5.14 Ground clearance 13
5.15 Required width of angled corridor 14
5.16 Required doorway entry depth 15
5.17 Required corridor width for side opening 16

6 Typical values and recommended limits for Technical dimensions 17
6.1 Reduced overall length 17
6.2 Overall height 18
6.3 Radial wheel deviation for mass group I, II, and III 18
6.4 Lateral wheel deviation for mass group I, II, and III 19
6.5 Radial handrim deviation for mass I, II, and III 19
6.6 Lateral handrim deviation for mass I, II, and III 20
6.7 Full occupied length 20
6.8 Reduced occupied length 21
6.9 Occupied width 22
6.10 Occupied height 23
6.11 Ramp transition angle 23
6.12 Wheelbase 24
6.13 Rear wheel track 26

6.14 Front wheel track 27
6.15 Camber 28
6.16 Toe of occupant mass group I, II, and III 29
6.17 Skew of occupant mass group I, II, and III 30
6.18 Castor rake of occupant mass group I, II, and III 31
6.19 Castor cant for occupant mass group I, II, and III 32
6.20 Castor trail for occupant mass group I, II, and III 32
6.21 Castor wheel misalignment of occupant mass group I, II, and III 33