ISO 11608-1:2014-12 (E)

Needle-based injection systems for medical use - Requirements and test methods - Part 1: Needle-based injection systems

Contents

Page

Foreword v

Introduction vii

1 Scope 1

2 Normative references 1

3 Termsanddefinitions 1

4 Symbols and abbreviated terms 3

5 Requirements 4
5.1 General 4
5.2 System designations 5
5.3 Risk analysis requirements 5
5.4 Uncertainty of measurement and conformance with specifications 5
5.5 General design requirements 5

6 Reagent and apparatus 7
6.1 General 7
6.2 Test liquid 7
6.3 Balance 7
6.4 Test surface for free-fall testing 7

7 Determination of dose accuracy 7
7.1 General 7
7.2 Dosing regions 8
7.3 Dose settings 9
7.3.1 Multi-dose containers (system designations A and C) 9
7.3.2 Single-dose containers (system designations B and D) 9
7.4 Assessment 9
7.4.1 General 9
7.4.2 Determination of dose accuracy limits 10
7.4.3 Determination of last-dose error and last-dose accuracy limits (system designations A and C) 11
7.4.4 Calculation of dose delivery efficiency (system designations B1 and D1, user-filled) 11
7.4.5 Calculation of tolerance intervals 12

8 Preparation and operation of NISs 12

9 Test matrix 13

10 Test descriptions 16
10.1 General 16
10.2 Cool, standard and warm atmosphere testing 16
10.2.1 Pre-conditioning 16
10.2.2 Testing 16
10.3 Last-dose testing (system designations A and C only) 17
10.3.1 General 17
10.3.2 Pre-conditioning 17

10.3.3 Testing 17
10.4 Life-cycle testing (systems designations A and B only) -- Pre-conditioning 17
10.5 Free-fall testing 17
10.6 Dry-heat and cold-storage testing -- Pre-conditioning 19
10.7 Damp-heat testing (system designations A and B only) -- Pre-conditioning 19
10.8 Cyclical testing (system designations A and B only) -- Pre-conditioning 19
10.9 Vibration testing -- Pre-conditioning 20
10.10 Electromagnetic compatibility (EMC) (systems with electronics only) 20
10.10.1 General 20
10.10.2 Exposure to electrostatic discharge -- Pre-conditioning 20
10.10.3 Radiated radio-frequency (RF) fields -- Pre-conditioning 20
10.10.4 Compliance criteria for electrostatic discharge 20
10.10.5 Radiated radio-frequency (RF) fields 21

11 Inspection 21
11.1 Visual inspection 21
11.2 Container inspection 21
11.3 Dose accuracy acceptance criteria 21

12 Test report 22

13 Information supplied by the manufacturer 22
13.1 General 22
13.2 Marking 22
13.2.1 General 22
13.2.2 Marking on the NIS 23
13.2.3 Marking on the user packaging 23
13.3 Instructions for use 23

Annex A (informative) Dose replicates, accuracy and testing rationale 25

Annex B (normative) One- and two-sided tolerance limit factors, k 29

Bibliography 40