DIN EN ISO 8537:2016-11 (E)

Sterile single-use syringes, with or without needle, for insulin (ISO 8537:2016)

Contents

Page

European foreword ... **4**

Annex ZA (informative) **Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC amended by Directive 2007/47/EEC** ... **6**

Foreword ... **8**

Introduction ... **9**

1 Scope ... **10**

2 Normative references ... **10**

3 Terms and definitions ... **11**

4 Types of syringes ... **14**

5 Requirements ... **14**
5.1 General requirements ... 14
5.2 Material selection ... 15
5.3 Colour coding ... 15
5.4 Extraneous matter ... 16
5.4.1 General ... 16
5.4.2 Limits for acidity or alkalinity ... 16
5.4.3 Limits for extractable metals ... 16
5.5 Lubrication ... 16
5.5.1 Lubrication of syringes ... 16
5.5.2 Lubrication of needle tube ... 17
5.6 Dimensions ... 17
5.6.1 Barrel and plunger stopper ... 17
5.6.2 Finger grips ... 17
5.7 Plunger/plunger stopper ... 17
5.7.1 General ... 17
5.7.2 Fit of plunger stopper in barrel ... 17
5.8 Nozzle ... 17
5.8.1 Conical fitting ... 17
5.8.2 Position of nozzle on end of barrel ... 17
5.9 Needle tubing and needles ... 18
5.9.1 Needles for syringe types 3 and 4 ... 18
5.9.2 Needle tubing for syringe types 5, 6, 7 and 8 ... 18
5.9.3 Bond between hub and needle tube ... 18
5.10 Standard test environmental conditions ... 18
5.11 Performance of assembled syringe ... 18
5.11.1 Dead space ... 18
5.11.2 Freedom from leakage at needle ... 19
5.11.3 Freedom from leakage past plunger stopper ... 19

6 Packaging ... **19**
6.1 Unit packaging and self-contained syringe units ... 19
6.1.1 General ... 19
6.1.2 Unit packaging providing sterile barrier syringes (types 1, 3, 5 and 7) ... 19
6.1.3 Self-contained syringes with sterile interiors (types 2, 4, 6 and 8) ... 19
6.2 Multiple-unit packaging (for syringe types 2, 4, 6 and 8) ... 20
6.3 User packaging ... 20

7 Information supplied by the manufacturer **20**
7.1 General 20
7.2 Syringes 20
7.2.1 General 20
7.2.2 Additional marking for self-contained syringes (syringe types 2, 4, 6 and 8) 21
7.3 Unit packaging (for syringe types 1, 3, 5 and 7) 21
7.4 Multiple unit packs (syringe types 2, 4, 6 and 8) 21
7.5 User packaging 22
7.6 Storage container 23
7.7 Transport wrapping 23

Annex A (normative) **Fluid for determination of acidity/alkalinity and extractable metals** **24**

Annex B (normative) **Test method for air leakage past syringe piston during aspiration and for separation of rubber stopper and plunger** **25**

Annex C (normative) **Test method for determination of forces required to operate piston** **27**

Annex D (normative) **Test method for determination of dead space** **29**

Annex E (normative) **Test method for liquid leakage at syringe piston and syringe nozzle/hub or needle/barrel unions during compression** **30**

Annex F (normative) **Test method for air leakage past nozzle/hub or needle/barrel unions during aspiration** **32**

Annex G (normative) **Preparation of extract for test for pyrogenicity and toxicity** **33**

Annex H (normative) **Syringe sizes and graduated scales** **34**

Bibliography **36**