

# DIN EN 80369-5:2017-10 (E)

Small-bore connectors for liquids and gases in healthcare applications - Part 5:  
Connectors for limb cuff inflation applications (IEC 80369-5:2016 + Cor. 1:2017)  
(includes Corrigendum :2017)

---

<b>Contents</b>	<b>Page</b>
European foreword .....	4
Annex ZA (normative) Normative references to international publications with their corresponding European publications .....	6
Annex ZZ (informative) Relationship between this European standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered .....	7
Introduction .....	8
1 * Scope .....	9
2 Normative references .....	9
3 Terms and definitions .....	10
4 General requirements .....	11
4.1 General requirements for the limb cuff inflation APPLICATION .....	11
4.2 Materials used for SMALL-BORE CONNECTORS .....	11
4.3 TYPE TESTS .....	12
5 Dimensional requirements for sphygmomanometer and cuff SMALL-BORE CONNECTORS .....	12
5.1 * Requirements for adult or paediatric PATIENT SMALL-BORE CONNECTORS (S1) .....	12
5.2 Void .....	12
6 Performance requirements .....	12
6.1 Air leakage .....	12
6.2 * Resistance to separation from axial load .....	12
Annex A (informative) Rationale and guidance .....	13
A.1 General guidance .....	13
A.2 Rationale for particular clauses and subclauses .....	13
Annex B (normative) SMALL-BORE CONNECTORS for the limb cuff inflation APPLICATION .....	15
Annex C (normative) Reference CONNECTORS .....	19
C.1 General requirements for reference CONNECTORS .....	19
C.2 * Sphygmomanometer and cuff S1 reference CONNECTORS .....	19
Annex D (informative) Assessment of MEDICAL DEVICES and their attributes with CONNECTIONS within this APPLICATION .....	20
Annex E (informative) Summary of the usability requirements for SMALL-BORE CONNECTORS for limb cuff inflation APPLICATIONS .....	21
E.1 USER PROFILE .....	21

E.2	Use scenarios .....	21
E.3	Use environments and scenarios .....	22
E.4	Generic USER needs .....	22
Annex F (informative) Summary of SMALL-BORE CONNECTOR design requirements for limb cuff inflation APPLICATIONS .....		23
Annex G (informative) Summary of assessment of the design of the CONNECTORS for limb cuff inflation APPLICATION .....		31
G.1	General .....	31
G.2	Summary of the engineering analysis of the design .....	31
G.2.1	NON-INTERCONNECTABLE analysis .....	31
G.2.2	S1 male to N1 male .....	31
G.3	Summary of the design VERIFICATION .....	31
G.4	Summary of the design validation .....	32
G.5	Summary of the design review .....	32
Annex H (informative) Obsolete limb cuff inflation CONNECTOR .....		33
Annex I (informative) Air leakage by pressure decay TEST METHOD .....		36
I.1	Principle .....	36
I.2	* Test conditions .....	36
I.2.1	Test sample preconditioning .....	36
I.2.2	Environmental test conditions .....	36
I.3	Apparatus .....	36
I.4	PROCEDURE .....	36
I.5	Test report .....	37
Annex J (informative) Resistance to separation from axial load TEST METHOD .....		38
J.1	Principle .....	38
J.2	* Test conditions .....	38
J.2.1	Test sample preconditioning .....	38
J.2.2	Environmental test conditions .....	38
J.3	Apparatus .....	38
J.4	PROCEDURE .....	38
J.5	Test report .....	38
Annex K (informative) Reference to the essential principles .....		39
Index of defined terms .....		41
Bibliography .....		42