

DIN EN ISO 80369-7:2017-10 (E)

Small-bore connectors for liquids and gases in healthcare applications - Part 7:
Connectors for intravascular or hypodermic applications (ISO 80369-7:2016,
Corrected version 2016-12-01)

Contents	Page
European foreword	3
Annex ZA (informative) Relationship between this European standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered	5
Foreword	7
Introduction	9
1 * Scope	10
2 Normative references	10
3 Terms and definitions	11
4 General requirements	12
4.1 General requirements for LUER CONNECTORS	12
4.2 Material used for LUER CONNECTORS	13
4.3 Type tests	13
5 * Dimensional requirements for LUER CONNECTORS	13
6 Performance requirements	13
6.1 Fluid leakage	13
6.1.1 Fluid leakage requirement	13
6.1.2 Leakage by pressure decay	13
6.1.3 Positive pressure liquid leakage	13
6.2 Sub-atmospheric pressure air leakage	14
6.3 Stress cracking	14
6.4 Resistance to separation from axial load	14
6.5 Resistance to separation from unscrewing	14
6.6 Resistance to overriding	14
Annex A (informative) Rationale and guidance	15
Annex B (normative) LUER CONNECTORS	19
Annex C (normative) Reference CONNECTORS	30
Annex D (informative) Assessment of MEDICAL DEVICES and their attributes with CONNECTIONS within this APPLICATION	37
Annex E (informative) Summary of the usability requirements for LUER CONNECTORS for intravascular or hypodermic APPLICATIONS	39
Annex F (informative) Summary of LUER CONNECTOR design requirements for intravascular or hypodermic APPLICATIONS	43
Annex G (informative) Summary of assessment of the design of the LUER CONNECTOR for intravascular or hypodermic APPLICATIONS	46
Annex H (informative) Reference to the essential principles	49
Annex I (informative) Terminology — Alphabetized index of defined terms	50
Bibliography	51