ISO 16840-2:2018-01 (E)

Wheelchair seating - Part 2: Determination of physical and mechanical characteristics of seat cushions intended to manage tissue integrity

Contents

Page

Foreword v

Introduction vi

1 Scope 1

2 Normative references 1

3 Termsanddefinitions 1

4 Symbols and abbreviated terms 2

5 Apparatus 2

6 Preparation of test cushion 7
6.1 Choice of cushion 7
6.2 Preconditioning the cushion 7
6.3 Setup 8

7 Sequence of testing 8

8 Frictional properties 8
8.1 Rationale 8
8.2 Test method 8
8.3 Test report 8

9 Impact damping under normal loading conditions 8
9.1 Rationale 8
9.2 Test method 8
9.3 Method of calculation 10
9.4 Test report 10

10 Recovery 11
10.1 Rationale 11
10.2 Test method 11
10.3 Test report 12

11 Loadedcontourdepthandoverloaddeflection 12
11.1 Rationale 12
11.2 Test method 12
11.3 Method of calculation 14
11.4 Test report 14

12 Water spillage 14
12.1 Rationale 14
12.2 Test method 14
12.3 Test report 14

13 Biocompatibility 14
13.1 Rationale 14
13.2 Test method 15

13.3 Test report 15

14 Hysteresis test 15
14.1 Rationale 15
14.2 Test method 15
14.3 Method of calculation 15
14.4 Test report 16

15 Test report 16

16 Disclosure requirement 16

Annex A (normative) Tapered uniform geometry RCLI 17

Annex C (informative) Horizontal stiffness and sliding resistance 22

Annex D (informative) RCLIs for testing of alternatively sized cushions 26

Bibliography 27