DIN EN ISO 11070:2019-04 (E)

Sterile single-use intravascular introducers, dilators and guidewires (ISO 11070:2014 + Amd 1:2018) (inclu des Amendment :2018)

Contents

Page

European foreword ... 4

European Foreword to Amendment A1 ... 5

Foreword to Amendment 1 ... 6

Introduction ... 7

1 Scope ... 8

2 Normative references ... 8

3 Terms and definitions ... 8

4 General requirements ... 12
4.1 Sterilization ... 12
4.2 Biocompatibility ... 12
4.3 Surface ... 12
4.4 Corrosion resistance ... 12
4.5 Radio-detectability ... 12
4.6 Information to be supplied by the manufacturer ... 12

5 Additional requirements for introducer needles ... 13
5.1 General ... 13
5.2 Size designation ... 13
5.3 Needle point ... 13
5.4 Hub ... 13
5.5 Information to be supplied by the manufacturer ... 13

6 Additional requirements for introducer catheters ... 13
6.1 General ... 13
6.2 Tip ... 14
6.3 Peak tensile force ... 14
6.4 Hub ... 14
6.5 Size designation ... 14
6.6 Information to be supplied by the manufacturer ... 15

7 Additional requirements for sheath introducers ... 15
7.1 General ... 15
7.2 Size designation ... 15
7.3 Freedom from leakage from sheath introducer ... 15
7.4 Freedom from leakage through haemostasis valve ... 15
7.5 Hub ... 15
7.6 Peak tensile force ... 15
7.7 Information to be supplied by the manufacturer ... 15

8 Additional requirements for guidewires ... 15
8.1 General ... 15
8.2 Size designation ... 16
8.3 Safety wire ... 16
8.4 Fracture test ... 16
8.5 Flexing test ... 16

8.6 Peak tensile force of guidewire 16
8.7 Information to be supplied by the manufacturer 17
9 Additional requirements for dilators 17
9.1 General 17
9.2 Size designation 17
9.3 Hub 17
9.4 Information to be supplied by the manufacturer 17
10 Additional requirements for kits containing combinations of devices specified in this International Standard 17
Annex A (informative) Guidance on materials and design 19
Annex B (normative) Test method for corrosion resistance 20
Annex C (normative) Method for determining peak tensile force of introducer catheters, sheath introducers, and dilators 21
Annex D (normative) Test method for liquid leakage from sheath introducers under pressure 23
Annex E (normative) Test method for liquid leakage through haemostasis valves of sheath introducers 25
Annex F (normative) Test method for fracture of guidewires 26
Annex G (normative) Test method for resistance of guidewires to damage by flexing 28
Annex H (normative) Method for determining peak tensile force of guidewires 30
Annex I (normative) Determination of strength of union of needle hub and needle 31
Bibliography 32