

DIN EN ISO 20697:2019-04 (E)

Sterile drainage catheters and accessory devices for single use (ISO 20697:2018, Corrected version 2018-09)

Contents		Page
EN ISO 20697:2018 (E) European foreword		4
Foreword		5
Introduction		6
1	Scope	7
2	Normative references	7
3	Terms and definitions	7
4	Intended performance	9
5	General requirements	9
5.1	Risk management	9
5.2	Biocompatibility	10
5.3	Detectability	10
5.4	Surface finish	10
5.5	Size designation	10
5.5.1	General	10
5.5.2	Outer diameter	10
5.5.3	Effective length	10
5.5.4	Nominal balloon inflation volume	11
5.6	Connector	11
5.7	MRI compatibility	12
5.8	Sterilization	12
6	Specific requirements	12
6.1	Kink stability	12
6.2	Corrosion resistance	12
6.3	Resistance to deformation	12
6.4	Peak tensile force	12
6.4.1	Connections	12
6.4.2	Drainage catheters and other accessory devices	13
6.5	Impact resistance	13
6.6	Flow rate	13
6.7	Retention strength	13
6.8	Balloon safety	13
6.9	Drainage catheter inflation lumen integrity and volume maintenance	14
6.9.1	General	14
6.9.2	Compliant balloon	14
6.9.3	Non-compliant balloon	14
6.10	Inflated balloon resistance to traction	14
6.11	Freedom from leakage during aspiration or vacuum	14
7	Information supplied by the manufacturer	15
7.1	General	15
7.2	Marking on the device and/or packaging	15
7.3	Instructions for use	15

Annex A (informative) Test method for determining kink stability	17
Annex B (normative) Test method for corrosion resistance.....	19
Annex C (normative) Test method for resistance to deformation by suction.....	20
Annex D (normative) Test method for determining peak tensile force of connections.....	21
Annex E (normative) Test method for determining peak tensile force of drainage catheter	22
Annex F (normative) Test method for impact resistance of collection device	24
Annex G (normative) Test method for determination of flow rate through drainage catheter	26
Annex H (informative) Test method for retention strength.....	28
Annex I (normative) Test method for determining balloon safety	30
Annex J (normative) Test method for determining inflation lumen leakage and/or function and/or balloon deflation (drainage catheter with compliant balloon).....	33
Annex K (normative) Test method for determining balloon size and deflation reliability (drainage catheter with non-compliant balloon)	35
Annex L (normative) Test method for determining inflated balloon resistance to traction.....	37
Annex M (normative) Test method for resistance to leakage during aspiration or vacuum	41
Bibliography.....	42