

# ISO 21366:2019 (E)

## Traditional Chinese medicine — General requirements for smokeless moxibustion devices

---

### Contents

	Foreword
	Introduction
1	Scope
2	Normative references
3	Terms and definitions
4	Test methods for smokeless moxibustion devices
4.1	General
4.2	Smoke density test
4.3	Smoke generation test
5	Safety requirements
5.1	Noxious gas
5.1.1	General
5.1.2	Carbon monoxide (CO)
5.1.3	Nitrogen dioxide (NO <sub>2</sub> )
5.1.4	Total volatile organic compounds (tVOCs)
5.1.5	Aldehyde
5.2	Moxibustion temperature
5.3	Device materials
6	Packaging and labelling
6.1	General
6.2	Smokeless statement
6.3	Protection of products
Annex A	(informative) WHO guidelines for indoor air quality
A.1	General
A.2	Particulate matter (PM)
A.2.1	PM <sub>10</sub>
A.2.2	PM <sub>2.5</sub>
A.3	Volatile organic compounds (VOCs)
A.3.1	Benzene
A.3.2	Styrene
A.3.3	Toluene
A.4	Formaldehyde
A.5	Carbon monoxide (CO)
A.6	Nitrogen dioxides (NO <sub>2</sub> )
Annex B	(normative) Test method for moxibustion smoke density measurement
B.1	General
B.2	Measuring apparatus for smoke density
B.2.1	General
B.2.2	Smoke chamber
B.2.3	Photometer
B.3	Test conditions
B.4	Formulae for calculating smoke density
B.4.1	General
B.4.2	Smoke density

- B.4.3** Maximum smoke density
- B.4.4** Correction factor of smoke density
- B.4.5** Correction value of smoke density
- B.4.6** Data selection
- B.5** Limit values of moxibustion smoke density
- B.5.1** General
- B.5.2** Limit value

**Annex C (normative) Test method for moxibustion smoke generation**

- C.1** General
- C.2** Measuring apparatus for smoke generation
  - C.2.1** General
  - C.2.2** Smoke chamber
    - C.2.2.1** General
    - C.2.2.2** Smoke area
    - C.2.2.3** CO2 area
    - C.2.2.4** Mesh board shell
    - C.2.2.5** Smoke detector
  - C.2.3** Fire alarm system
- C.3** Test conditions
  - C.3.1** Sample preparation
  - C.3.2** Sample selection
  - C.3.3** Test environment
  - C.3.4** Calibration
  - C.3.5** Test interval
- C.4** Limit values of moxibustion smoke generation

Page count: 14