

# ISO 3826-1:2019 (E)

## Plastics collapsible containers for human blood and blood components — Part 1: Conventional containers

---

### Contents

	Foreword
	Introduction
1	Scope
2	Normative references
3	Terms and definitions
4	Dimensions
5	Design
5.1	General
5.2	Air content
5.3	Emptying under pressure
5.4	Pilot samples
5.5	Rate of collection
5.6	Collection and transfer tube(s)
5.7	Blood-taking needle
5.8	Outlet port(s)
5.9	Suspension
6	Requirements
6.1	General
6.2	Physical requirements
6.2.1	Conditions of manufacture
6.2.2	Sterilization
6.2.3	Transparency
6.2.4	Coloration
6.2.5	Thermal stability
6.2.6	Water vapour transmission
6.2.7	Resistance to leakage
6.2.8	Particulate contamination
6.3	Chemical requirements
6.3.1	Requirements for the raw container or sheeting
6.3.2	Requirements for the test fluid
6.4	Biological requirements
6.4.1	General
6.4.2	Impermeability for microorganisms
6.4.3	Compatibility
7	Packaging
8	Labelling
8.1	General
8.2	Label on plastics container
8.3	Label on over-package
8.4	Label on shipping box
8.5	Label requirements
9	Anticoagulant and/or preservative solution

## **Annex A (normative) Chemical tests**

- A.1 General**
- A.2 Determination of residue on ignition**
- A.3 Preparation of the test fluid**
- A.4 Tests**
  - A.4.1 Determination of oxidizable constituents**
  - A.4.2 Determination of ammonia**
  - A.4.3 Determination of chloride ions**
  - A.4.4 Determination of metals**
    - A.4.4.1 Heavy metals related to Pb<sup>2+</sup>**
    - A.4.4.2 Alternative methods for testing for heavy metals**
  - A.4.5 Determination of acidity or alkalinity**
  - A.4.6 Determination of the evaporation residue**
  - A.4.7 Determination of turbidity and degree of opalescence**
    - A.4.7.1 General**
    - A.4.7.2 Reagents**
      - A.4.7.2.1 Hydrazine sulfate solution**
      - A.4.7.2.2 Hexamethylenetetramine solution**
      - A.4.7.2.3 Primary opalescent suspension**
      - A.4.7.2.4 Standard of opalescence**
      - A.4.7.2.5 Reference suspensions**
    - A.4.7.3 Expression of results**
  - A.4.8 Determination of degree of coloration**
    - A.4.8.1 General**
    - A.4.8.2 Method 1**
    - A.4.8.3 Method 2**
    - A.4.8.4 Expression of results**
  - A.4.9 Determination of the UV absorption**
  - A.4.10 Determination of plasticizer as extractable di(2-ethylhexyl) phthalate (DEHP)**
    - A.4.10.1 Reagents**
    - A.4.10.2 Preparation of standard solutions**
      - A.4.10.2.1 Solution 1**
      - A.4.10.2.2 Solution 2**
      - A.4.10.2.3 Standard solutions A to E**
    - A.4.10.3 Calibration curves**
    - A.4.10.4 Extraction procedure**
    - A.4.10.5 Expression of results**

## **Annex B (normative) Physical tests**

- B.1 Transparency test**
- B.2 Test for rate of collection**
- B.3 Test for permanence of labelling**
- B.4 Test for particulate contamination**
- B.5 Test for sterile connection of tubing**

## **Annex C (normative) Biological tests**

- C.1 General**
- C.2 Preparation of the test solutions**
  - C.2.1 Test fluid I (polar extractant)**
  - C.2.2 Test fluid II (non-polar extractant)**
- C.3 Test for impermeability to microorganisms**
- C.4 Test for bacterial endotoxins**
- C.5 Test for cytotoxicity**
- C.6 Test for haemolysis**
  - C.6.1 General**
  - C.6.2 Preparation of the erythrocyte suspension**
  - C.6.3 Procedure**
- C.7 Biological test methods**