

# ISO 8536-12:2021 (E)

## Infusion equipment for medical use — Part 12: Check valves for single use

---

### Contents

	Foreword
1	Scope
2	Normative references
3	Terms and definitions
4	Designation
5	Materials
6	Physical requirements
6.1	Particulate contamination
6.2	Leakage
6.3	ISO 80369 (all parts) compatibility
6.4	Counterflow pressure resistance
6.5	Flow rate
6.6	Blocking performance
6.7	Opening pressure
6.8	Protective caps
7	Chemical requirements
8	Biological requirements
8.1	General
8.2	Sterility
8.3	Pyrogenicity
9	Labelling
9.1	General
9.2	Label on unit container
9.3	Label on shelf or multi-unit container
10	Packaging
11	Disposal
Annex A	(normative) Physical tests
A.1	General
A.2	Test for leakage
A.3	Test for pressure resistance in counterflow direction
A.4	Determination of flow rate
A.5	Test for blocking performance
A.6	Determination of opening pressure
A.6.1	Opening pressure when first used
A.6.2	Opening pressure after closure of the check valve