

# ISO 18666:2021-11 (E)

## Traditional Chinese medicine - General requirements of moxibustion devices

---

<b>Contents</b>		<b>Page</b>
Foreword .....		iv
Introduction .....		v
<b>1</b>	<b>Scope .....</b>	<b>1</b>
<b>2</b>	<b>Normative references .....</b>	<b>1</b>
<b>3</b>	<b>Terms and definitions .....</b>	<b>1</b>
<b>4</b>	<b>Configuration .....</b>	<b>2</b>
<b>4.1</b>	<b>Structure .....</b>	<b>2</b>
<b>4.1.1</b>	<b>Size and shape .....</b>	<b>2</b>
<b>4.1.2</b>	<b>Visual inspection .....</b>	<b>2</b>
<b>4.2</b>	<b>Expression of dimensions .....</b>	<b>3</b>
<b>5</b>	<b>Materials .....</b>	<b>3</b>
<b>5.1</b>	<b>General .....</b>	<b>3</b>
<b>5.2</b>	<b>Body of moxibustion devices .....</b>	<b>3</b>
<b>5.3</b>	<b>Moxibustion materials .....</b>	<b>3</b>
<b>5.3.1</b>	<b>Moxa floss .....</b>	<b>3</b>
<b>5.3.2</b>	<b>Moxa sticks .....</b>	<b>3</b>
<b>6</b>	<b>Moxibustion temperature .....</b>	<b>4</b>
<b>6.1</b>	<b>General .....</b>	<b>4</b>
<b>6.2</b>	<b>Moxibustion with tube .....</b>	<b>4</b>
<b>6.3</b>	<b>Moxibustion devices other than moxibustion with tube .....</b>	<b>4</b>
<b>6.4</b>	<b>Moxibustion devices for repeated use .....</b>	<b>4</b>
<b>7</b>	<b>Safety requirements .....</b>	<b>4</b>
<b>7.1</b>	<b>General .....</b>	<b>4</b>
<b>7.2</b>	<b>Environmental protection .....</b>	<b>4</b>
<b>8</b>	<b>Packaging .....</b>	<b>5</b>
<b>8.1</b>	<b>Primary packaging .....</b>	<b>5</b>
<b>8.2</b>	<b>Secondary packaging .....</b>	<b>5</b>
<b>9</b>	<b>Labelling .....</b>	<b>5</b>
<b>9.1</b>	<b>Primary label .....</b>	<b>5</b>
<b>9.2</b>	<b>Secondary label .....</b>	<b>5</b>
<b>10</b>	<b>Transport and storage .....</b>	<b>6</b>
<b>11</b>	<b>Instructions for use .....</b>	<b>6</b>
<b>Annex A (informative)</b>	<b>Test methods to determine the moxibustion temperature .....</b>	<b>7</b>
<b>Annex B (informative)</b>	<b>Method of artificial drying of mugwort leaves by heat .....</b>	<b>10</b>
<b>Bibliography .....</b>		<b>11</b>