

DIN EN ISO 11137-2:2023-08 (E)

Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose (ISO 11137-2:2013 + Amd 1:2022) (includes Amendment A1:2023)

| Contents | | Page |
|-----------------|---|-------------|
| | European foreword | 4 |
| | European foreword to Amendment | 5 |
| | Foreword to Amendment | 6 |
| | Introduction | 7 |
| 1 | Scope | 8 |
| 2 | Normative references | 8 |
| 3 | Terms, definitions, and abbreviated terms | 8 |
| 3.1 | Terms and definitions | 8 |
| 3.2 | Abbreviated terms | 10 |
| 4 | Definition and maintenance of product families for dose setting, dose substantiation, and sterilization dose auditing | 11 |
| 4.1 | General | 11 |
| 4.2 | Defining product families | 11 |
| 4.3 | Designation of product to represent a product family for performance of a verification dose experiment or sterilization dose audit | 12 |
| 4.4 | Maintaining product families | 13 |
| 4.5 | Effect of failure of establishment of sterilization dose or of a sterilization dose audit on a product family | 14 |
| 5 | Selection and testing of product for establishing the sterilization dose | 14 |
| 5.1 | Nature of product | 14 |
| 5.2 | Sample item portion (SIP) | 15 |
| 5.3 | Manner of sampling | 16 |
| 5.4 | Microbiological testing | 16 |
| 5.5 | Irradiation | 16 |
| 6 | Methods of dose establishment | 16 |
| 7 | Method 1: dose setting using bioburden information | 17 |
| 7.1 | Rationale | 17 |
| 7.2 | Procedure for Method 1 for product with an average bioburden greater than or equal to 1,0 for multiple production batches | 18 |
| 7.3 | Procedure for Method 1 for product with an average bioburden greater than or equal to 1,0 for a single production batch | 24 |
| 7.4 | Procedure for Method 1 for product with an average bioburden in the range 0,1 to 0,9 for multiple or single production batches | 26 |
| 8 | Method 2: Dose setting using fraction positive information from incremental dosing to determine an extrapolation factor | 27 |
| 8.1 | Rationale | 27 |
| 8.2 | Procedure for Method 2A | 28 |
| 8.3 | Procedure for Method 2B | 31 |
| 9 | Method VD_{max} -- Substantiation of a selected sterilization dose | 35 |
| 9.1 | Selected doses and rationale | 35 |
| 9.2 | Procedure for Method VD_{max25} for multiple production batches | 36 |
| 9.3 | Procedure for Method VD_{max25} for a single production batch | 41 |
| 9.4 | Procedure for Method VD_{max15} for multiple production batches | 44 |
| 9.5 | Procedure for Method VD_{max15} for a single production batch | 47 |

| | | |
|---|---|-----------|
| 10 | Sterilization dose audit | 50 |
| 10.1 | Purpose and frequency | 50 |
| 10.2 | Procedure for auditing a sterilization dose established using Method 1, Method 2A, or Method 2B | 50 |
| 10.3 | Procedure for auditing a sterilization dose substantiated using Method VDmax25 or Method VDmax15 | 53 |
| 10.4 | Failure of a sterilization dose audit | 59 |
| 11 | Worked examples | 59 |
| 11.1 | Worked examples for Method 1 | 59 |
| 11.2 | Worked examples for Method 2 | 61 |
| 11.3 | Worked examples for Method VDmax | 69 |
| 11.4 | Worked example of a sterilization dose audit for a dose established using Method 1, the findings from which necessitated augmentation of the sterilization dose | 71 |
| 11.5 | Worked example of a sterilization dose audit for a dose established using Method 2A, the findings from which necessitated augmentation of the sterilization dose | 71 |
| 11.6 | Worked example of a sterilization dose audit for a sterilization dose substantiated using Method VDmax25 | 72 |
| Annex ZA (informative) Relationship between this European Standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered | | 74 |
| Annex ZB (informative) Relationship between this European Standard and the General Safety and Performance Requirements of Regulation (EU) 2017/746 aimed to be covered | | 77 |
| Bibliography | | 79 |