ISO 24571:2022-06 (E)

Traditional Chinese medicine - General requirements for the basic safety and essential performance of electro-acupuncture stimulators

Contents

Page

Foreword ..... **v**

Introduction ..... **vi**

1 Scope **1**

2 Normative references **1**

3 Terms and definitions **1**

4 General requirements **2**
4.1 General 2

5 General requirements for testing EA stimulators **2**

6 Classification of EA stimulators in medical electrical (ME) equipment and ME systems **2**
6.1 General 2
6.2 Protection against electric shock 2
6.3 Mode of operation 2

7 Identification, marking and documents of EA stimulators **2**
7.1 General 2
7.2 Electrical input power from the supply mains 3
7.3 Instructions for use 3
7.3.1 General 3
7.3.2 Additional information 3

8 Protection against electrical hazards from EA stimulators **3**
8.1 General 3
8.2 Classification of applied parts 3

9 Protection against mechanical hazards of EA stimulators **3**
9.1 General 3
9.2 Disconnection prevention 4

10 Protection against unwanted and excessive radiation hazards **4**

11 Protection against excessive temperatures and other hazards **4**

12 Accuracy of controls and instruments and protection against hazardous outputs **4**
12.1 Accuracy of controls and instruments 4
12.1.1 General 4
12.1.2 Output amplitude 4
12.1.3 Pulse parameters 4
12.1.4 Independent control of multiple output channels 4
12.2 Usability of ME equipment 5
12.3 Alarm systems 5
12.4 Protection against hazardous output 5
12.4.1 General 5
12.4.2 Supply voltage fluctuations 5
12.4.3 Output interlock 5
12.4.4 Output indicator 5
12.4.5 Limitation of output parameters 5

13 Hazardous situations and fault conditions **6**

14 Programmable electrical medical systems (PEMS) **6**

15 Construction of ME equipment **6**
15.1 General 6
15.2 Needle clips and leads 6
16 ME systems **6**
17 Electromagnetic compatibility of ME equipment and ME system **6**
18 Electromagnetic compatibility – requirements and tests **6**
18.1 General 6
18.2 Emissions tests 6
18.2.1 General 6
18.2.2 Patient lead 6
18.3 Immunity 6
18.3.1 General 6
18.3.2 EA stimulator with acupuncture needle 7
Annex A (informative) **Configuration of EA stimulator system** **8**
Annex B (normative) **Test methods for EA stimulator** **9**
Annex C (informative) **The safety issue of excessive total current per session** **10**
Annex D (informative) **Test methods for EA stimulator for various impedance conditions** **11**
Bibliography **12**