

# ISO 24571:2022-06 (E)

## Traditional Chinese medicine - General requirements for the basic safety and essential performance of electro-acupuncture stimulators

---

<b>Contents</b>	<b>Page</b>
Foreword.....	v
Introduction.....	vi
<b>1 Scope.....</b>	<b>1</b>
<b>2 Normative references.....</b>	<b>1</b>
<b>3 Terms and definitions.....</b>	<b>1</b>
<b>4 General requirements.....</b>	<b>2</b>
4.1 General.....	2
<b>5 General requirements for testing EA stimulators.....</b>	<b>2</b>
<b>6 Classification of EA stimulators in medical electrical (ME) equipment and ME systems.....</b>	<b>2</b>
6.1 General.....	2
6.2 Protection against electric shock.....	2
6.3 Mode of operation.....	2
<b>7 Identification, marking and documents of EA stimulators.....</b>	<b>2</b>
7.1 General.....	2
7.2 Electrical input power from the supply mains.....	3
7.3 Instructions for use.....	3
7.3.1 General.....	3
7.3.2 Additional information.....	3
<b>8 Protection against electrical hazards from EA stimulators.....</b>	<b>3</b>
8.1 General.....	3
8.2 Classification of applied parts.....	3
<b>9 Protection against mechanical hazards of EA stimulators.....</b>	<b>3</b>
9.1 General.....	3
9.2 Disconnection prevention.....	4
<b>10 Protection against unwanted and excessive radiation hazards.....</b>	<b>4</b>
<b>11 Protection against excessive temperatures and other hazards.....</b>	<b>4</b>
<b>12 Accuracy of controls and instruments and protection against hazardous outputs.....</b>	<b>4</b>
12.1 Accuracy of controls and instruments.....	4
12.1.1 General.....	4
12.1.2 Output amplitude.....	4
12.1.3 Pulse parameters.....	4
12.1.4 Independent control of multiple output channels.....	4
12.2 Usability of ME equipment.....	5
12.3 Alarm systems.....	5
12.4 Protection against hazardous output.....	5
12.4.1 General.....	5
12.4.2 Supply voltage fluctuations.....	5
12.4.3 Output interlock.....	5
12.4.4 Output indicator.....	5
12.4.5 Limitation of output parameters.....	5
<b>13 Hazardous situations and fault conditions.....</b>	<b>6</b>
<b>14 Programmable electrical medical systems (PEMS).....</b>	<b>6</b>

<b>15</b>	<b>Construction of ME equipment</b> .....	<b>6</b>
15.1	General.....	6
15.2	Needle clips and leads.....	6
<b>16</b>	<b>ME systems</b> .....	<b>6</b>
<b>17</b>	<b>Electromagnetic compatibility of ME equipment and ME system</b> .....	<b>6</b>
<b>18</b>	<b>Electromagnetic compatibility - requirements and tests</b> .....	<b>6</b>
18.1	General.....	6
18.2	Emissions tests.....	6
	18.2.1 General.....	6
	18.2.2 Patient lead.....	6
18.3	Immunity.....	6
	18.3.1 General.....	6
	18.3.2 EA stimulator with acupuncture needle.....	7
	<b>Annex A (informative) Configuration of EA stimulator system</b> .....	<b>8</b>
	<b>Annex B (normative) Test methods for EA stimulator</b> .....	<b>9</b>
	<b>Annex C (informative) The safety issue of excessive total current per session</b> .....	<b>10</b>
	<b>Annex D (informative) Test methods for EA stimulator for various impedance conditions</b> .....	<b>11</b>
	<b>Bibliography</b> .....	<b>12</b>