ISO 22441:2022-08 (E)

Sterilization of health care products - Low temperature vaporized hydrogen peroxide - Requirements for the development, validation and routine control of a sterilization process for medical devices

Contents

Page

Foreword v

Introduction vi

1 Scope 1
1.1 Inclusions 1
1.2 Exclusions 1

2 Normative references 2

3 Termsanddefinitions 2

4 Quality management system elements 10

5 Sterilizing agent characterization 11
5.1 General 11
5.2 Sterilant and sterilizing agent 11
5.3 Microbicidal effectiveness 11
5.4 Effects on materials 12
5.5 Safety and the environment 12

6 Process and equipment characterization 13
6.1 General 13
6.2 Process characterization 13
6.3 Equipment characterization 14

7 Productdefinition 15

8 Processdefinition 16

9 Validation 17
9.1 General 17
9.2 Installation qualification (IQ) 18
9.2.1 General 18
9.2.2 Equipment 18
9.3 Operational qualification (OQ) 19
9.4 Performance qualification (PQ) 19
9.5 Review and approval of validation 21

10 Routine monitoring and control 21

11 Product release from sterilization 22

12 Maintaining process effectiveness 22
12.1 General 22
12.2 Recalibration 22
12.3 Maintenance of equipment 23
12.4 Requalification 23
12.5 Assessment of change 23

Annex A (normative) Factors to be considered in selection of microorganisms for demonstrating microbicidal effectiveness 24

Annex B (normative) Approach 1 -- Process definition based on inactivation of the microbial population in its natural state (bioburden method) 26

Annex C (normative)Approach2--Processdefinitionbasedoninactivationofreference microorganisms and knowledge of bioburden (BI/bioburden method) 28

Annex D (normative)Approach3--Conservativeprocessdefinitionbasedoninactivation of reference microorganisms (overkill method) 30

Annex E (informative) Guidance on application of this document 33

Annex F (informative) Schematic example of a VH2O2 sterilization cycle 56

Annex G (informative) Environmental aspects 57

Annex H (informative)Justificationforthenumberoftemperaturesensorsandbiological indicators (BIs) 62

Annex I (informative) Aspects of VH2O2 64

Annex J (informative) Recommended validation tests 67

Annex K (informative) Recommended validation test procedures 72

Bibliography 78