

ISO 13004:2022-10 (E)

Sterilization of health care products - Radiation - Substantiation of selected sterilization dose: Method VD<(Index)max><(hoch)SD>

Contents		Page
Foreword		v
Introduction		vi
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	Definition and maintenance of product families for sterilization dose substantiation and sterilization dose auditing	5
4.1	General	5
4.2	Defining product families	5
4.3	Designation of product to represent a product family	6
4.3.1	Product to represent a product family	6
4.3.2	Master product	7
4.3.3	Equivalent product	7
4.3.4	Simulated product	7
4.4	Maintaining product families	8
4.4.1	Periodic review	8
4.4.2	Modification to either product or manufacturing process, or both	8
4.4.3	Records	8
4.5	Consequence of failure of sterilization dose substantiation or sterilization dose audit	8
5	Selection and testing of product for substantiating and auditing a selected sterilization dose	8
5.1	Nature of product	8
5.2	Sample item portion (SIP)	9
5.3	Manner of sampling	10
5.4	Microbiological testing	11
5.5	Irradiation	11
6	Method VD SD max -- Substantiation of a selected sterilization dose of 17,5 kGy, 20 kGy, 22,5 kGy, 27,5 kGy, 30 kGy, 32,5 kGy or 35 kGy	12
6.1	Rationale	12
6.2	Procedure for Method VDmax SD for multiple production batches	12
6.2.1	General	12
6.2.2	Stage 1: Obtain samples of product	13
6.2.3	Stage 2: Determine average bioburden	13
6.2.4	Stage 3: Obtain the selected sterilization dose	13
6.2.5	Stage 4: Obtain VD SD max	14
6.2.6	Stage 5: Perform verification dose experiment	15
6.2.7	Stage 6: Interpretation of results	15
6.2.8	Confirmatory verification dose experiment	16
6.3	Procedure for Method VDmax SD for a single production batch	17
6.3.1	Rationale	17
6.3.2	General	17
6.3.3	Stage 1: Obtain samples of product	17
6.3.4	Stage 2: Determine average bioburden	18
6.3.5	Stage 3: Obtain the selected sterilization dose	18

6.3.6	Stage 4: Obtain VDmax SD	19
6.3.7	Stage 5: Perform verification dose experiment	19
6.3.8	Stage 6: Interpretation of results	19
6.3.9	Confirmatory verification dose experiment	20
7	Maintaining process effectiveness	21
7.1	General	21
7.2	Frequency of determination of bioburden	21
7.3	Sterilization dose audit	21
7.3.1	Frequency	21
7.3.2	Outcome	22
7.3.3	Procedure for auditing a sterilization dose substantiated using Method VDmax SD	22
7.3.4	Failure of a sterilization dose audit	25
8	Tables of values for SIP	26
9	Worked examples	51
9.1	Substantiation of a selected sterilization dose of 17,5 kGy (SIP less than 1,0)	51
9.2	Substantiation of a selected sterilization dose of 30 kGy (SIP equal to 1,0)	52
9.3	Sterilization dose audit for a sterilization dose substantiated using	52
9.4	Method VDmax 22,5	52
	Bibliography	54