

DIN EN ISO 13004:2023-10 (E)

Sterilization of health care products - Radiation - Substantiation of selected sterilization dose: Method VD_{\max}^{SD} (ISO 13004:2022)

Contents	Page
European foreword.....	4
Foreword.....	5
Introduction.....	6
1 Scope.....	7
2 Normative references.....	7
3 Terms and definitions.....	7
4 Definition and maintenance of product families for sterilization dose substantiation and sterilization dose auditing.....	11
4.1 General.....	11
4.2 Defining product families.....	11
4.3 Designation of product to represent a product family.....	12
4.3.1 Product to represent a product family.....	12
4.3.2 Master product.....	13
4.3.3 Equivalent product.....	13
4.3.4 Simulated product.....	13
4.4 Maintaining product families.....	14
4.4.1 Periodic review.....	14
4.4.2 Modification to either product or manufacturing process, or both.....	14
4.4.3 Records.....	14
4.5 Consequence of failure of sterilization dose substantiation or sterilization dose audit.....	14
5 Selection and testing of product for substantiating and auditing a selected sterilization dose.....	14
5.1 Nature of product.....	14
5.2 Sample item portion (SIP).....	15
5.3 Manner of sampling.....	16
5.4 Microbiological testing.....	17
5.5 Irradiation.....	17
6 Method VD_{\max}^{SD} — Substantiation of a selected sterilization dose of 17,5 kGy, 20 kGy, 22,5 kGy, 27,5 kGy, 30 kGy, 32,5 kGy or 35 kGy.....	18
6.1 Rationale.....	18
6.2 Procedure for Method VD_{\max}^{SD} for multiple production batches.....	18
6.2.1 General.....	18
6.2.2 Stage 1: Obtain samples of product.....	19
6.2.3 Stage 2: Determine average bioburden.....	19
6.2.4 Stage 3: Obtain the selected sterilization dose.....	19
6.2.5 Stage 4: Obtain VD_{\max}^{SD}	20
6.2.6 Stage 5: Perform verification dose experiment.....	21
6.2.7 Stage 6: Interpretation of results.....	21
6.2.8 Confirmatory verification dose experiment.....	22
6.3 Procedure for Method VD_{\max}^{SD} for a single production batch.....	23
6.3.1 Rationale.....	23
6.3.2 General.....	23
6.3.3 Stage 1: Obtain samples of product.....	23

6.3.4	Stage 2: Determine average bioburden	24
6.3.5	Stage 3: Obtain the selected sterilization dose	24
6.3.6	Stage 4: Obtain VD_{max}^{SD}	25
6.3.7	Stage 5: Perform verification dose experiment	25
6.3.8	Stage 6: Interpretation of results	25
6.3.9	Confirmatory verification dose experiment	26
7	Maintaining process effectiveness	27
7.1	General	27
7.2	Frequency of determination of bioburden	27
7.3	Sterilization dose audit	27
7.3.1	Frequency	27
7.3.2	Outcome	28
7.3.3	Procedure for auditing a sterilization dose substantiated using Method VD_{max}^{SD}	28
7.3.4	Failure of a sterilization dose audit	31
8	Tables of values for SIP	32
9	Worked examples	57
9.1	Substantiation of a selected sterilization dose of 17,5 kGy (SIP less than 1,0)	57
9.2	Substantiation of a selected sterilization dose of 30 kGy (SIP equal to 1,0)	58
9.3	Sterilization dose audit for a sterilization dose substantiated using	58
9.4	Method $VD_{max}^{22,5}$	58
	Bibliography	60