

DIN EN ISO 13004:2023-10 (E)

Sterilization of health care products - Radiation - Substantiation of selected sterilization dose: Method VD_{max}^{SD} (ISO 13004:2022)

Contents

	Page
European foreword	4
Foreword	5
Introduction	6
1 Scope	7
2 Normative references	7
3 Terms and definitions	7
4 Definition and maintenance of product families for sterilization dose substantiation and sterilization dose auditing	11
4.1 General	11
4.2 Defining product families	11
4.3 Designation of product to represent a product family	12
4.3.1 Product to represent a product family	12
4.3.2 Master product	13
4.3.3 Equivalent product	13
4.3.4 Simulated product	13
4.4 Maintaining product families	14
4.4.1 Periodic review	14
4.4.2 Modification to either product or manufacturing process, or both	14
4.4.3 Records	14
4.5 Consequence of failure of sterilization dose substantiation or sterilization dose audit	14
5 Selection and testing of product for substantiating and auditing a selected sterilization dose	14
5.1 Nature of product	14
5.2 Sample item portion (SIP)	15
5.3 Manner of sampling	16
5.4 Microbiological testing	17
5.5 Irradiation	17
6 Method VD_{max}^{SD} — Substantiation of a selected sterilization dose of 17,5 kGy, 20 kGy, 22,5 kGy, 27,5 kGy, 30 kGy, 32,5 kGy or 35 kGy	18
6.1 Rationale	18
6.2 Procedure for Method VD_{max}^{SD} for multiple production batches	18
6.2.1 General	18
6.2.2 Stage 1: Obtain samples of product	19
6.2.3 Stage 2: Determine average bioburden	19
6.2.4 Stage 3: Obtain the selected sterilization dose	19
6.2.5 Stage 4: Obtain VD_{max}^{SD}	20
6.2.6 Stage 5: Perform verification dose experiment	21
6.2.7 Stage 6: Interpretation of results	21
6.2.8 Confirmatory verification dose experiment	22
6.3 Procedure for Method VD_{max}^{SD} for a single production batch	23
6.3.1 Rationale	23
6.3.2 General	23
6.3.3 Stage 1: Obtain samples of product	23

6.3.4	Stage 2: Determine average bioburden	24
6.3.5	Stage 3: Obtain the selected sterilization dose	24
6.3.6	Stage 4: Obtain VD_{max}^{SD}	25
6.3.7	Stage 5: Perform verification dose experiment	25
6.3.8	Stage 6: Interpretation of results	25
6.3.9	Confirmatory verification dose experiment	26
7	Maintaining process effectiveness	27
7.1	General	27
7.2	Frequency of determination of bioburden	27
7.3	Sterilization dose audit	27
7.3.1	Frequency	27
7.3.2	Outcome	28
7.3.3	Procedure for auditing a sterilization dose substantiated using Method VD_{max}^{SD}	28
7.3.4	Failure of a sterilization dose audit	31
8	Tables of values for SIP	32
9	Worked examples	57
9.1	Substantiation of a selected sterilization dose of 17,5 kGy (SIP less than 1,0)	57
9.2	Substantiation of a selected sterilization dose of 30 kGy (SIP equal to 1,0)	58
9.3	Sterilization dose audit for a sterilization dose substantiated using	58
9.4	Method $VD_{max}^{22,5}$	58
Bibliography	60	