

# ISO 10555-4:2023-11 (E)

## Intravascular catheters - Sterile and single-use catheters - Part 4: Balloon dilatation catheters

---

<b>Contents</b>		<b>Page</b>
Foreword .....		iv
<b>1</b>	<b>Scope .....</b>	<b>1</b>
<b>2</b>	<b>Normative references .....</b>	<b>1</b>
<b>3</b>	<b>Terms and definitions .....</b>	<b>1</b>
<b>4</b>	<b>Requirements .....</b>	<b>2</b>
4.1	General .....	2
4.2	Detectability of the balloon position .....	2
4.3	Designation of nominal size .....	2
4.4	Physical requirements .....	2
4.4.1	Balloon rated burst pressure (RBP) .....	2
4.4.2	Balloon fatigue; freedom from leakage and damage on inflation .....	2
4.4.3	Balloon deflation time .....	2
4.4.4	Balloon diameter to inflation pressure (balloon compliance) .....	2
4.4.5	Crossing profile .....	3
4.4.6	Balloon removal .....	3
4.5	Information to be supplied with the catheter .....	3
Annex A (normative) Test for rated burst pressure (RBP) .....		4
Annex B (normative) Balloon fatigue test for freedom from leakage and damage on inflation .....		6
Annex C (normative) Test for balloon deflation time .....		8
Annex D (normative) Test for balloon diameter to inflation pressure (balloon compliance) .....		10
Annex E (normative) Determination of crossing profile .....		12
Annex F (normative) Test method for balloon removal .....		14
Annex G (informative) Rationale and guidance .....		16
Bibliography .....		17