

# ISO 10555-1:2023-11 (E)

## Intravascular catheters - Sterile and single-use catheters - Part 1: General requirements

---

<b>Contents</b>		<b>Page</b>
Foreword .....		v
1	Scope .....	1
2	Normative references .....	1
3	Terms and definitions .....	1
4	Requirements .....	5
4.1	Risk approach .....	5
4.2	Usability engineering .....	5
4.3	Sterilization .....	5
4.4	Shelf life .....	5
4.5	Detectability .....	5
4.6	Biocompatibility .....	5
4.7	Surface .....	5
4.8	Corrosion resistance .....	6
4.9	Peak tensile force .....	6
4.10	Freedom from leakage during pressurization .....	7
4.11	Freedom from leakage during aspiration .....	7
4.12	Hubs .....	7
4.13	Flowrate .....	8
4.14	Power injection burst pressure .....	8
4.15	Packaging system .....	8
4.16	Simulated use, kink and/or torque testing to consider depending on device design, intended use, and risk analysis .....	8
4.17	Coating integrity and/or particulate testing to consider depending on device design, intended use, and risk analysis .....	9
4.18	Distal tip stiffness testing to consider for neurovascular applications .....	9
5	Designation of nominal size .....	9
5.1	Nominal outside diameter .....	9
5.2	Nominal inside diameter .....	9
5.3	Nominal effective length .....	9
6	Information to be supplied with the catheter .....	10
6.1	General .....	10
6.2	Marking on the device and/or primary packaging .....	10
6.3	Instructions for use .....	11
6.4	Marking on the secondary packaging .....	11
Annex A (normative)	Test method for corrosion resistance .....	12
Annex B (normative)	Method for determining peak tensile force .....	13
Annex C (normative)	Test method for liquid leakage under pressure .....	16
Annex D (normative)	Test method for air leakage into hub assembly during aspiration .....	18
Annex E (normative)	Determination of flow rate through catheter .....	20

<b>Annex F (normative) Test for burst pressure under static conditions .....</b>	<b>22</b>
<b>Annex G (normative) Power injection tests for flowrate and device pressure (only for products indicated for power injection) .....</b>	<b>25</b>
<b>Annex H (informative) Units of measurement systems other than those specified in this document .....</b>	<b>30</b>
<b>Annex I (normative) Test method for air leakage under water .....</b>	<b>32</b>
<b>Annex J (informative) Rationale and guidance .....</b>	<b>34</b>
<b>Annex K (informative) Test methods for distal tip stiffness for neurovascular applications .....</b>	<b>41</b>
<b>Bibliography .....</b>	<b>43</b>