

# DIN EN ISO 10555-4:2024-03 (E)

## Intravascular catheters - Sterile and single-use catheters - Part 4: Balloon dilatation catheters (ISO 10555-4:2023)

---

<b>Contents</b>	<b>Page</b>
<b>European foreword</b> .....	<b>3</b>
<b>Foreword</b> .....	<b>4</b>
<b>1 Scope</b> .....	<b>5</b>
<b>2 Normative references</b> .....	<b>5</b>
<b>3 Terms and definitions</b> .....	<b>5</b>
<b>4 Requirements</b> .....	<b>6</b>
4.1 General.....	6
4.2 Detectability of the balloon position .....	6
4.3 Designation of nominal size.....	6
4.4 Physical requirements.....	6
4.4.1 Balloon rated burst pressure (RBP).....	6
4.4.2 Balloon fatigue; freedom from leakage and damage on inflation.....	6
4.4.3 Balloon deflation time.....	6
4.4.4 Balloon diameter to inflation pressure (balloon compliance).....	6
4.4.5 Crossing profile.....	7
4.4.6 Balloon removal.....	7
4.5 Information to be supplied with the catheter.....	7
<b>Annex A (normative) Test for rated burst pressure (RBP)</b> .....	<b>8</b>
<b>Annex B (normative) Balloon fatigue test for freedom from leakage and damage on inflation</b> .....	<b>10</b>
<b>Annex C (normative) Test for balloon deflation time</b> .....	<b>12</b>
<b>Annex D (normative) Test for balloon diameter to inflation pressure (balloon compliance)</b> .....	<b>14</b>
<b>Annex E (normative) Determination of crossing profile</b> .....	<b>16</b>
<b>Annex F (normative) Test method for balloon removal</b> .....	<b>18</b>
<b>Annex G (informative) Rationale and guidance</b> .....	<b>20</b>
<b>Bibliography</b> .....	<b>21</b>