

# DIN EN ISO 15883-2:2025-07 (E)

## Washer-disinfectors - Part 2: Requirements and tests for washer-disinfectors employing thermal disinfection for critical and semi-critical medical devices (ISO 15883-2:2024)

---

<b>Contents</b>	<b>Page</b>
<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Performance requirements</b> .....	<b>3</b>
4.1 General.....	3
4.2 Cleaning.....	4
4.3 Disinfecting.....	4
4.4 Temperature of inner surfaces of processed devices.....	5
4.5 Water quality.....	5
<b>5 Mechanical and control requirements</b> .....	<b>5</b>
5.1 Lumen and powered devices.....	5
5.1.1 Irrigation.....	5
5.1.2 Verification of flow through lumen and powered devices.....	6
5.2 Control systems.....	6
5.3 Process verification.....	6
<b>6 Testing for conformity</b> .....	<b>6</b>
6.1 General.....	6
6.2 Tests for soil removal from chamber walls, load carrier(s) and load.....	7
6.3 Thermometric tests.....	7
6.3.1 General.....	7
6.3.2 Temperature of outer surfaces of devices.....	7
6.3.3 Temperature of inner surfaces of devices.....	8
6.4 Pressure and flow measurement.....	9
<b>7 Information to be provided for the WD</b> .....	<b>9</b>
<b>8 Information to be requested from the purchaser by the supplier of the WD</b> .....	<b>9</b>
<b>Annex A (informative) Summary of test programmes</b> .....	<b>10</b>
<b>Annex B (informative) Guidance on the designation of a medical device to a product family for cleaning and thermal disinfection processes</b> .....	<b>11</b>
<b>Bibliography</b> .....	<b>14</b>