

# ISO 80601-2-74:2026-04 (E)

## Medical electrical equipment - Part 2-74: Particular requirements for basic safety and essential performance of respiratory humidifying equipment

---

### Contents

Page

Foreword.....	v
Introduction.....	vi
201.1 Scope, object and related standards.....	1
201.2 Normative references .....	3
201.3 Terms and definitions.....	5
201.4 General requirements.....	20
201.5 General requirements for testing of <i>ME equipment</i> .....	23
201.6 Classification of <i>ME equipment</i> and <i>ME systems</i> .....	25
201.7 <i>ME equipment</i> identification, marking and documents .....	26
201.8 Protection against electrical hazards from <i>ME equipment</i> .....	33
201.9 Protection against mechanical hazards of <i>ME equipment</i> and <i>ME systems</i> .....	33
201.10 Protection against unwanted and excessive radiation hazards.....	35
201.11 Protection against excessive temperatures and other hazards .....	35
201.12 Accuracy of controls and instruments and protection against hazardous outputs .....	38
201.13 Hazardous situations and fault conditions for <i>ME Equipment</i> .....	43
201.14 Programmable electrical medical systems (PEMS) .....	44
201.15 Construction of <i>ME equipment</i> .....	45
201.16 <i>ME systems</i> .....	45
201.16.2 Accompanying documents of an <i>ME system</i> .....	45
201.17 Electromagnetic compatibility of <i>ME equipment</i> and <i>ME systems</i> .....	46
201.101 Breathing system connectors and ports.....	46
201.102 Requirements for the breathing system and accessories .....	49
201.103 Liquid container.....	49
201.104 Functional connection.....	50
202 Electromagnetic disturbances — Requirements and tests .....	50
206 Usability .....	51
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems .....	53
211 Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment.....	53
Annex C (informative) Guide to marking and labelling requirements for <i>ME equipment</i> and <i>ME systems</i> .....	54
Annex D (informative) Symbols on marking.....	60
Annex AA (informative) Particular guidance and rationale.....	62
Annex BB (normative) Determination of the accuracy of the displayed measured gas temperature .....	82
Annex CC (normative) Determination of the humidification output.....	84

<b>Annex DD (normative) Specific enthalpy calculations .....</b>	<b>89</b>
<b>Annex EE (normative) Removable temperature sensors and mating ports.....</b>	<b>91</b>
<b>Annex FF (normative) Reference temperature sensor.....</b>	<b>94</b>
<b>Annex GG (informative) Saturation vapour pressure.....</b>	<b>97</b>
<b>Annex HH (informative) Liquid fill port .....</b>	<b>98</b>
<b>Annex II (informative) Reference to the IMDRF <i>essential principles</i> and labelling guidances.....</b>	<b>101</b>
<b>Annex JJ (informative) Terminology — Alphabetized index of defined terms .....</b>	<b>105</b>
<b>Bibliography .....</b>	<b>110</b>