

ISO 24996:2026-09 (E)

Traditional Chinese medicine - General requirements and recommendations for the basic safety and essential performance of the transcutaneous electrical acupoint stimulators

Contents	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
Terms and definitions	1
4 General requirements	2
5 General requirements for testing TEAS	2
6 Classification of TEAS in medical electrical (ME) equipment and ME systems	2
7 Identification, marking and documents of TEAS	3
7.1 General	3
7.2 Marking on the outside of TEAS or TEAS parts	3
7.3 Marking of controls and instruments	3
7.4 Accompanying documents	3
8 Protection against electrical hazards from TEAS	3
9 Protection against mechanical hazards from TEAS in ME equipment and ME systems	3
10 Protection against unwanted and excessive radiation hazards	3
11 Protection against excessive temperatures and other hazards	3
12 Accuracy of controls and instruments and protection against hazardous outputs	4
12.1 General	4
12.2 Accuracy of controls and instruments	4
12.2.1 Pulse width range	4
12.2.2 Frequency range	4
12.2.3 Intensity range	4
12.2.4 Current density of the electrode contact area	4
12.3 Protection against hazardous outputs	4
12.3.1 General	4
12.3.2 Supply voltage fluctuations	5
12.3.3 Output interlock	5
12.3.4 Output indicator	5
12.3.5 Safety criteria for stimulation signals	5
13 Hazardous situations and fault conditions	5
14 Programmable electrical medical systems (PEMS)	5
15 Construction of ME equipment	5

15.1	General	5
15.2	Electrodes and clips	5
15.2.1	General	5
15.2.2	The area of the electrode	5
15.2.3	Auricular electrodes and clips	6
15.2.4	The adhesive force of the electrode	7
15.3	Lead	7
15.4	Operation of electrodes in fault circuit condition	7
16	ME systems	7
17	Electromagnetic compatibility of ME equipment and ME systems	7
Annex A (informative)RationaleforkeyoutputparameterlimitsofTEAS		8
Bibliography		9