ISO 9919 :2005-03 (E)

Medical electrical equipment_- Particular requirements for the basic safety and essential performance of pulse oximeter equipment for medical use

Contents

Page

Foreword ... vii

Introduction ... viii

1 Scope ... 1

2 Normative references ... 1

3 Terms and definitions ... 2

4 General requirements and requirements for tests ... 7
4.101 Other test methods ... 7
4.102 Acceptance criteria ... 8
4.103 Pulse oximeter equipment, parts and accessories ... 8

5 Classification ... 8

6 Identification, marking and documents ... 8
6.1 Marking on the outside of equipment or equipment parts ... 8
6.8.1 General ... 9
6.8.2 Instructions for use ... 9
6.8.3 Technical description ... 11

7 Power input ... 11

8 Basic safety categories ... 11

9 Removable protective means ... 11

10 Environmental conditions ... 12
10.1 Transport and storage ... 12

11 Not used ... 12

12 Not used ... 12

13 General ... 12

14 Requirements related to classification ... 12
14.6 Types B, BF and CF equipment ... 12

15 Limitation of voltage and/or energy ... 12

16 Enclosures and protective covers ... 12

17 Separation ... 12

18 Protective earthing, functional earthing and potential equalization ... 12

19 Continuous leakage currents and patient auxiliary currents ... 13
19.4 Tests ... 13

20 Dielectric strength ... 13
20.4 Tests ... 13

21 * Mechanical strength ... 13
21.5 13
21.101 * Shock and vibration ... 13
21.102 * Shock and vibration for transport ... 14

22 Moving parts ... 15

23 Surfaces, corners and edges ... 15

24 Stability in normal use ... 15

25 Expelled parts 15
26 Vibration and noise 16
27 Pneumatic and hydraulic power 16
28 Suspended masses 16
29 X-Radiation 16
30 Alpha, beta, gamma, neutron radiation and other particle radiation 16
31 Microwave radiation 16
32 Light radiation (including lasers) 16
33 Infra-red radiation 16
34 Ultraviolet radiation 16
35 Acoustical energy (including ultrasonics) 16
36 * Electromagnetic compatibility 17
37 Locations and basic requirements 17
38 Marking, accompanying documents 17
39 Common requirements for category AP and category APG equipment 17
40 Requirements and tests for category AP equipment, parts and components thereof 17
41 Requirements and tests for category APG equipment, parts and components thereof 17
42 Excessive temperatures 18
43 Fire prevention 18
43.101 * Pulse oximeter equipment used in conjunction with oxidants 18
43.101.1 Ignitable material 18
43.101.2 Sparking 19
44 Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization, disinfection and compatibility 19
44.6 * Ingress of liquids 19
44.7 Cleaning, sterilization and disinfection 19
45 Pressure vessels and parts subject to pressure 19
46 Human errors 20
47 Electrostatic charges 20
48 Biocompatibility 20
49 Interruption of the power supply 20
49.101 Power-failure alarm condition 20
49.102 Pulse oximeter equipment operation following interruption of the power supply 20
49.102.1 Settings and data storage following short interruptions or automatic switchover 20
49.102.2 Operation following long interruptions 20
50 Accuracy of operating data 21
50.101 * SpO_2 accuracy of pulse oximeter equipment 21
50.101.1 * Specification 21
50.101.2 Determination of SpO_2 accuracy 21
50.102 Accuracy under conditions of motion 22
50.103 Accuracy under conditions of low perfusion 22
50.104 Pulse rate accuracy 23
51 Protection against hazardous output 23
51.101 * Data update period 23
51.102 Detection of pulse oximeter probe and probe cable extender fault 23

52 Abnormal operation and fault-conditions 23
53 Environmental tests 24
54 General 24
55 Enclosures and covers 24
56 Components and general assembly 24
57 Mains parts, components and layout 24
58 Protective earthing — Terminals and connections 24
59 Construction and layout 24
101 * Signal inadequacy 24
102 * Pulse oximeter probes and probe cable extenders 25
102.1 General 25
102.2 Labelling 25
103 Saturation pulse information signal 25
104 Alarm systems 25
201.1.2 * Assignment of priority 25
201.5.4 * Default alarm preset 26
201.8 Alarm signal inactivation states 26
201.8.3 Indication and access 26
105 Appendices of IEC 60601-1:1988 26
Annex AA (informative) **Rationale** 27
Annex BB (informative) **Skin temperature at the pulse oximeter probe** 38
Annex CC (informative) **Determination of accuracy** 42
Annex DD (informative) **Calibration standards** 50
Annex EE (informative) **Guideline for evaluating and documenting SpO_2 accuracy in human subjects** 51
Annex FF (informative) **Simulators, calibrators and functional testers for pulse oximeter equipment** 58
Annex GG (informative) **Concepts of equipment response time** 68
Annex HH (informative) **Reference to the Essential Principles** 72
Annex II (informative) **Environmental aspects** 74
Annex JJ (informative) **Index of defined terms** 76
Bibliography 78

Tables

Table AA.1 — Qualitative assessment of pulse oximeter equipment shock and vibration environment 28
Table AA.2 — Allowable maximum temperatures for skin contact with medical electrical equipment applied parts (adapted from Table 22, IEC/CDV 60601-1:2004) 30
Table BB.1 — Pulse oximeter probe safe application time and source 40
Table EE.1 — Example of target plateaus and ranges 54
Table HH.1 — Correspondence between this International Standard and the Essential Principles 72
Table II.1 — Environmental aspects addressed by clauses of this International Standard 75

Figures

Figure CC.1 — Synthesized calibration data (base case) 43

Figure CC.2 — Constant offset has been added to base case 44

Figure CC.3 — Tilt has been added to base case 45

Figure CC.4 — Graphical representation for the definition of local bias (Test sensor SpO_2 as a function of reference S_R) 46

Figure CC.5 — Graphical representation for the definition of local bias and mean bias (Test sensor SpO_2 as a function of reference S_R) 46

Figure EE.1 — Example of desaturation-time profile 54

Figure FF.1 — Sample calibration curve for pulse oximeter equipment 60

Figure FF.2 — Interface of a functional tester that uses a photodiode and LED to interact with a pulse oximeter probe 61

Figure FF.3 — Interface of a functional tester that uses a dye mixture 62

Figure FF.4 — Interface of a functional tester that uses a liquid crystal modulator 63

Figure FF.5 — Absorbency of blue bandage material (measured in reflection) used in a special test pulse oximeter probe with great patient-to-patient variability of calibration 65

Figure FF.6 — Calibration of high-variability pulse oximeter probe in controlled desaturation study on five test subjects 66

Figure FF.6 — Calibration of high-variability pulse oximeter probe in controlled desaturation study on five test subjects (*continued*) 67

Figure GG.1 — Illustration of fidelity of pulse oximeter equipment performance in tracking saturation changes 68

Figure GG.2 — Illustration of effect of different averaging times on fidelity 69

Figure GG.3 — Graphic representation of components of alarm system delay 70

Figure GG.4 — Illustration of the effects of different averaging times on a more rapid and noisier desaturation signal 71