ISO 11137-2:2006-04 (E)

Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose

Contents

Page

Foreword v

Introduction vi

1 Scope 1

2 Normative references 1

3 Abbreviations, terms and definitions 1
3.1 Abbreviations 1
3.2 Terms 3

4 Definition and maintenance of product families for dose setting, dose substantiation and sterilization dose auditing 4
4.1 General 4
4.2 Defining product families 4
4.3 Designation of product to represent a product family for performance of a verification dose experiment or sterilization dose audit 5
4.4 Maintaining product families 6
4.5 Effect of failure of establishment of sterilization dose or of a sterilization dose audit on a product family 7

5 Selection and testing of product for establishing and verifying the sterilization dose 7
5.1 Nature of product 7
5.2 Sample item portion (SIP) 8
5.3 Manner of sampling 8
5.4 Microbiological testing 9
5.5 Irradiation 9

6 Methods of dose establishment 9

7 Method 1: dose setting using bioburden information 10
7.1 Rationale 10
7.2 Procedure for Method 1 for product with an average bioburden W 1,0 for multiple production batches 11
7.3 Procedure for Method 1 for product with an average bioburden W 1,0 for a single production batch 16
7.4 Procedure for Method 1 for product with an average bioburden in the range 0,1 to 0,9 for multiple or single production batches 18

8 Method 2: Dose setting using fraction positive information from incremental dosing to determine an extrapolation factor 18
8.1 Rationale 18
8.2 Procedure for Method 2A 19
8.3 Procedure for Method 2B 22

9 Method VDmax -- Substantiation of 25 kGy or 15 kGy as the sterilization dose 25
9.1 Rationale 25
9.2 Procedure for Method VDmax 25 for multiple production batches 26
9.3 Procedure for Method VDmax 25 for a single production batch 29
9.4 Procedure for Method VDmax 15 for multiple production batches 30

9.5 Procedure for Method VDmax 15 for a single production batch 33

10 Auditing sterilization dose 34
10.1 Purpose and frequency 34
10.2 Procedure for auditing a sterilization dose established using Method 1 or Method 2 35
10.3 Procedure for auditing a sterilization dose substantiated using VDmax 37

11 Worked examples 41
11.1 Worked examples for Method 1 41
11.2 Worked examples for Method 2 44
11.3 Worked examples for Method VDmax 53
11.4 Worked example of a sterilization dose audit for a dose established using Method 1, the findings from which necessitated augmentation of the sterilization dose 55
11.5 Worked example of a sterilization dose audit for a dose established using Method 2A, the findings from which necessitated augmentation of the sterilization dose 56
11.6 Worked example of a sterilization dose audit for a sterilization dose substantiated using Method VDmax 25 58

Bibliography 59