

ISO 17665-1:2006-08 (E)

Sterilization of health care products - Moist heat - Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices

Contents		Page
Foreword		v
Introduction		vi
1	Scope	1
1.1	Inclusions	1
1.2	Exclusions	1
2	Normative references	2
3	Terms and definitions	3
4	Quality management system elements	10
4.1	Documentation	10
4.2	Management responsibility	10
4.3	Product realization	10
4.4	Measurement, analysis and improvement -- Control of non-conforming product	10
5	Sterilizing agent characterization	11
5.1	Sterilizing agent	11
5.2	Microbicidal effectiveness	11
5.3	Materials effects	11
5.4	Environmental consideration	11
6	Process and equipment characterization	11
6.1	Process	11
6.1.1	General	11
6.1.2	Saturated steam processes	12
6.1.3	Contained product processes	12
6.2	Equipment	13
7	Product definition	14
8	Process definition	15
9	Validation	17
9.1	General	17
9.2	Installation qualification (IQ)	17
9.2.1	Equipment	17
9.2.2	Installation	17
9.2.3	Function	17
9.3	Operational qualification (OQ)	18
9.4	Performance qualification (PQ)	18
9.5	Review and approval of validation	19
10	Routine monitoring and control	20
11	Product release from sterilization	21

12	Maintaining process effectiveness	21
12.1	Demonstration of continued effectiveness	21
12.2	Recalibration	21
12.3	Maintenance of equipment	21
12.4	Requalification	22
12.5	Assessment of change	22
Annex A (informative) Guidance		23
Annex B (informative) Process definition based on inactivation of the microbial population in its natural state (bioburden-based method)		27
Annex C (informative) Process definition based on the inactivation of a reference microorganism and a knowledge of bioburden on product items to be sterilized (combined bioburden/biological indicator based method)		28
Annex D (informative) Conservative process definition based on inactivation of reference microorganisms (overkill method)		29
Annex E (informative) Operating cycles		31
Bibliography		36