

ISO 17593:2007-04 (E)

Clinical laboratory testing and in vitro medical devices - Requirements for in vitro monitoring systems for self-testing of oral anticoagulant therapy

Contents		Page
Foreword		v
Introduction		vi
1	Scope	1
2	Normative references	1
3	Terms and definitions	2
4	Design and development	8
4.1	General requirements	8
4.2	Measuring interval	8
4.3	Safety	8
4.4	Risk management	8
4.4.1	Acceptability of risks	8
4.4.2	Risk assessment	8
4.5	Ergonomic and human factor aspects	9
4.6	Quality assurance and risk controls	9
4.6.1	General	9
4.6.2	Measurement verification	10
4.6.3	Control of system performance	10
4.6.4	Verification of self-testing performance	10
4.6.5	Evaluation of user compliance in following the manufacturer's and the physician's instructions	10
4.7	Metrological traceability	10
5	Information supplied by the manufacturer	11
5.1	General requirements	11
5.2	Labels for the oral-anticoagulation monitoring instrument	11
5.3	Instructions for use of the oral-anticoagulation monitoring system	12
5.4	Labels for the reagents and control material	13
5.5	Instructions for use for reagents and control material	13
6	Safety and reliability testing	14
6.1	General requirements	14
6.1.1	Protocol	14
6.1.2	Instruments and reagents	15
6.1.3	Acceptance criteria	15
6.2	Protection against electric shock	15
6.3	Protection against mechanical hazards	15
6.4	Electromagnetic compatibility	15
6.5	Resistance to heat	15
6.6	Resistance to moisture and liquids	15
6.7	Protection against liberated gases, explosion and implosion	15
6.8	Instrument components	15
6.9	Performance test	15
6.10	Mechanical resistance to shock, vibration and impact	16
6.10.1	Vibration test protocol	16
6.10.2	Drop test protocol	16
6.11	Temperature exposure limits	16

6.11.1	High-temperature test protocol	16
6.11.2	Low-temperature test protocol	17
6.12	Humidity-exposure test protocol	17
6.13	Reagent storage and use testing	17
7	Training and education programs	17
7.1	Training of healthcare providers	17
7.2	Education of patients and other users	18
8	System performance verification	19
8.1	General	19
8.2	Contributors to measurement uncertainty	19
8.3	System performance verification study	19
8.4	Verification of measurement precision	20
8.4.1	General	20
8.4.2	Verification of measurement repeatability	20
8.4.3	Verification of intermediate measurement precision	21
8.4.4	Data analysis	22
8.5	Verification of system accuracy	24
8.5.1	General requirements	24
8.5.2	Study population	24
8.5.3	Samples	25
8.5.4	Instruments and reagents	25
8.5.5	Manufacturer's selected measurement procedure	26
8.5.6	Study design	26
8.5.7	Procedure	27
8.5.8	Data analysis	28
8.6	Minimum acceptable system accuracy	31
8.6.1	System accuracy requirement	31
8.6.2	System accuracy assessment	31
8.6.3	Data presentation	32
9	User performance evaluation	32
9.1	General	32
9.2	Study sites	32
9.3	Subjects	33
9.4	Instruments and materials	33
9.5	Evaluation of user proficiency	33
9.6	Acceptance criteria and data assessment	34
9.7	Evaluation of instructions for use	34
Annex A (normative) Additional requirements for electromagnetic compatibility		35
Annex B (informative) Traceability chain examples		37
Annex C (informative) Sample size calculation to estimate bias ([42] in the Bibliography)		40
Annex D (informative) Example of an uncertainty calculation for a prothrombin INR determination using an oral anticoagulation monitoring system		41
Annex E (informative) Elements of quality assurance of oral-anticoagulation monitoring systems ..		45
Annex F (informative) Application of performance criteria to published evaluations of oral- anticoagulation monitoring systems		46
Bibliography		51