

ISO 11135-1:2007-05 (E)

Sterilization of health care products - Ethylene oxide - Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices

Contents		Page
Foreword		iv
Introduction		v
1	Scope	1
2	Normative references	2
3	Terms and definitions	2
4	Quality management systems	9
4.1	Documentation	9
4.2	Management responsibility	9
4.3	Product realization	9
4.4	Measurement, analysis and improvement -- Control of nonconforming product	10
5	Sterilizing agent characterization	10
5.1	Sterilizing agent	10
5.2	Microbicidal effectiveness	10
5.3	Material effects	10
5.4	Environmental considerations	10
6	Process and equipment characterization	10
6.1	Process characterization	10
6.2	Equipment characterization	11
7	Product definition	12
7.1	General	12
7.2	Product safety and performance	12
7.3	Microbiological quality	12
7.4	Documentation	13
8	Process definition	13
9	Validation	14
9.1	Installation qualification	14
9.2	Operational qualification	14
9.3	Performance qualification	14
9.4	Varying load configurations	16
9.5	Review and approval of validation	16
10	Routine monitoring and control	18
11	Product release from sterilization	19
12	Maintaining process effectiveness	19
12.1	General	19
12.2	Maintenance of equipment	19
12.3	Requalification	19

12.4	Assessment of change	20
Annex A (normative)	Determination of lethal rate of the sterilization process -- Biological indicator/bioburden approach	21
Annex B (normative)	Conservative determination of lethal rate of the sterilization process -- Overkill approach	23
Annex C (informative)	General guidance	25
Bibliography	41