

DIN EN ISO 8041-1:2017-10 (E)

Human response to vibration - Measuring instrumentation - Part 1: General purpose vibration meters (ISO 8041-1:2017)

Contents	Page
European foreword	5
Foreword	6
Introduction	7
1 Scope	8
2 Normative references	8
3 Terms, definitions and symbols	9
3.1 Terms and definitions	9
3.1.1 General	9
3.1.2 Frequency-weighted values	11
3.2 Symbols	13
4 Reference environmental conditions	15
5 Performance specifications	15
5.1 General characteristics	15
5.2 Display of signal magnitude	17
5.2.1 General	17
5.2.2 Resolution and refresh rate	17
5.2.3 Stabilization, measurement start and display times	18
5.3 Electrical output	18
5.4 Vibration sensitivity	18
5.5 Accuracy of indication at reference frequency under reference conditions	19
5.6 Frequency weightings and frequency responses	19
5.6.1 Parameters	19
5.6.2 Band-limiting filter	20
5.6.3 a-v transition filter	20
5.6.4 Upward-step filter	21
5.6.5 Overall frequency weighting	21
5.6.6 Tolerances	21
5.7 Amplitude linearity	22
5.8 Instrument noise	23
5.9 Signal-burst response	23
5.10 Overload indication	26
5.11 Under-range indication	27
5.12 Time averaging	27
5.13 Running r.m.s. acceleration	27
5.14 Reset	28
5.15 Timing facilities	28
5.16 Electrical cross-talk	28
5.17 Vibration transducer characteristics	28
5.18 Power supply	28
6 Mounting	29
7 Environmental and electromagnetic criteria	29
7.1 General	29
7.2 Air temperature	29
7.3 Surface temperature	30
7.4 Electrostatic discharge	30
7.5 Radio-frequency emissions and public-power-supply disturbances	30
7.6 Immunity to AC power-frequency fields and radio-frequency fields	31
7.7 Ingress of water and dust	31
8 Provision for use with auxiliary devices	32
9 Instrument marking	32

10	Instrument documentation	32
11	Testing and calibration	32
12	Pattern evaluation	35
12.1	General	35
12.2	Testing requirements	35
12.3	Submission for testing	36
12.4	Marking of the vibration meter and information in the instrument documentation	36
12.5	Mandatory facilities and general requirements	36
12.6	Initial instrument preparation	36
12.7	Indication at the reference frequency under reference conditions	36
12.8	Electrical cross-talk	37
12.9	Vibration transducer	37
12.10	Amplitude linearity and under-range indication	38
12.10.1	Electrical tests of amplitude linearity	38
12.10.2	Mechanical tests of amplitude linearity	39
12.11	Frequency weightings and frequency responses	40
12.11.1	General	40
12.11.2	Mechanical tests of frequency response	40
12.11.3	Electrical tests of frequency response	41
12.11.4	Conformance	42
12.12	Instrument noise	42
12.13	Signal-burst response	43
12.14	Overload indication	43
12.15	Reset	43
12.16	Combined axis outputs	44
12.17	AC electrical output	44
12.18	Timing facilities	44
12.19	Power supply	44
12.20	Environmental, electrostatic and radio-frequency tests	44
12.20.1	General	44
12.20.2	Expanded uncertainties for measurements of environmental conditions	45
12.20.3	Acclimatization requirements for tests of the influence of air temperature and relative humidity	45
12.20.4	Test of the influence of air temperature and relative humidity combined	45
12.20.5	Influence of surface temperature	46
12.20.6	Influence of electrostatic discharges	46
12.20.7	Radio-frequency emissions and public-power-supply disturbances	46
12.20.8	Immunity to AC power-frequency fields and radio-frequency fields	47
12.21	Test report	48
13	Validation of one-off instruments	49
13.1	General	49
13.2	Testing requirements	49
13.3	Test object	49
13.4	Submission for testing	50
13.5	Marking of the one-off instrument and information in the instrument documentation	50
13.6	Mandatory facilities and general requirements	50
13.7	Initial instrument preparation	50
13.8	Test procedure	50
13.9	Indication at the reference frequency under reference conditions	51
13.10	Test parameters	52
13.10.1	Vibration measurement chain for hand-arm vibration	52
13.10.2	Vibration measurement chain for whole-body vibration	52
13.10.3	Vibration measurement chain low-frequency whole-body vibration	53
13.11	Conducting the test	53
13.12	Instrument noise	54
13.13	Electrical cross-talk	54
13.14	Overload indication	54

13.15	Timing facilities	54
13.16	Test report	55
14	Periodic verification	55
14.1	General	55
14.2	Testing requirements	55
14.3	Test object	55
14.4	Submission for testing	55
14.5	Preliminary inspection	56
14.6	Marking of the vibration meter and information in the instrument documentation	56
14.7	Test procedure	56
14.8	Test parameters	57
14.8.1	Vibration measurement chain for hand-arm vibration	57
14.8.2	Vibration measurement chain for whole-body vibration	57
14.8.3	Vibration measurement chain low-frequency whole-body vibration	57
14.9	Conducting the test	58
14.10	Test report	58
15	<i>In situ</i> checks	58
15.1	General	58
15.2	Preliminary inspection	59
15.3	Vibration sensitivity (field calibration)	59
Annex A (normative) Specification for field vibration calibrator		60
Annex B (informative) Frequency weightings		62
Annex C (informative) Realization of frequency weighting filters		81
Annex D (informative) Running r.m.s. time averaging		85
Annex E (informative) Vibration transducer characteristics		88
Annex F (informative) Tests for mounting systems		91
Annex G (normative) Instrument documentation		94
Annex H (normative) Phase response requirements for measurement of non-r.m.s. quantities		99
Annex I (informative) Guidelines for the estimation of the instrumental measurement uncertainty		107
Bibliography		113