

DIN 6868-157:2014-11 (E)

Image quality assurance in diagnostic X-ray departments - Part 157: X-ray ordinance acceptance and constancy test of image display systems in their environment

Contents	Page
Foreword	4
1 Scope	6
2 Normative references	6
3 Terms and definitions	7
4 Symbols and abbreviations	13
5 Test conditions.....	14
6 Test equipment.....	15
6.1 General.....	15
6.2 Luminance meter.....	15
6.2.1 Luminance meter for the acceptance test.....	15
6.2.2 Luminance meter for the constancy test.....	15
6.3 Illuminance meter / luxmeter	15
6.4 Test images.....	16
7 Room classes	17
8 Acceptance test for image display systems with diagnostic quality	19
8.1 Scope of testing.....	19
8.2 Visual tests.....	20
8.2.1 General	20
8.2.2 Overall image quality	20
8.2.3 Greyscale resolution of the IMAGE DISPLAY DEVICE.....	22
8.2.4 LUMINANCE uniformity (visual)	23
8.2.5 Colour impression and uniformity.....	23
8.2.6 Pixel errors	24
8.2.7 Image geometry.....	24
8.2.8 Test based on clinical reference images	25
8.2.9 Additional tests for mammography image display systems.....	25
8.3 Metrological tests.....	25
8.3.1 Minimum luminance	25
8.3.2 Maximum luminance.....	25
8.3.3 Ambient luminance.....	25
8.3.4 Illuminance	25
8.3.5 Maximum luminance ratio	25
8.3.6 Uniformity of the image display device	26
8.3.7 Uniformity of multiple image display devices	27
8.3.8 Luminance response curve	27
8.4 Minimum requirements (without mammography)	28
8.5 Additional requirements for image display systems used for mammography	30
8.6 Mobile systems.....	31
9 Reference values for the constancy test.....	31
10 Constancy test for image display systems with diagnostic quality	31

11	Significant changes to an image display system.....	33
12	Image display systems with review quality	33
Annex A	(normative) Methods of measuring luminance	34
A.1	General	34
A.2	Measurement methods.....	34
A.2.1	Method A: Telescopic method.....	34
A.2.2	Method B: Close-range luminance meter in combination with a luxmeter	35
A.2.3	Method C: Front-integrated luminance meter in combination with a luxmeter	36
A.2.4	Method D: Back-integrated luminance meter in combination with an illuminance meter/luxmeter	37
A.3	Notes on the measuring methods.....	37
Annex B	(informative) Typical errors.....	38
Annex C	(informative) Test report templates.....	39
C.1	Example of a test report template for the acceptance test.....	39
C.2	Example of a test report template for the constancy test.....	50
C.3	Example of a test report template for the daily constancy test.....	52
	Bibliography.....	53
	Keyword index.....	55

Figures

Figure 1	— Test images for use on image display systems (images a) through f) according to DIN EN 62563-1:2014-01, Table 2, and images g) through l) according to AAPM).....	16
Figure 2	— Overall image quality test using test image TG18-OIQ.....	22
Figure 3	— Enlarged view of test image TG18-MP, depicting the 8-bit and 10-bit markers	23
Figure 4	— Positions of the measuring points for the luminance uniformity test.....	26
Figure A.1	— Method A - Telescopic method	34
Figure A.2	— Method B - Close-range luminance meter in combination with a luxmeter.....	35
Figure A.3	— Method C - Front-integrated luminance meter in combination with luxmeter	36
Figure A.4	— Method D - Back-integrated luminance meter in combination with a luxmeter	37

Tables

Table 1	— Definition of room classes with their respective requirements.....	18
Table 2	— Visual tests for acceptance testing.....	19
Table 3	— Metrological tests for acceptance testing.....	20
Table 4	— Minimum requirements for image display devices.....	28
Table 5	— Additional requirements for image display systems used in mammography and stereotactic mammography.....	30
Table 6	— Visual tests for the constancy test.....	32
Table 7	— Metrological test items in the constancy test.....	32