

ISO/ASTM 51818:2020-06 (E)

Practice for dosimetry in an electron beam facility for radiation processing at energies between 80 and 300 keV

Contents	Page
1 Scope	1
2 Referenced documents	2
3 Terminology	2
4 Significance and use	3
5 Selection and calibration of the dosimetry system	3
6 Installation and operational qualification	4
7 Performance qualification	5
8 Routine process control	5
9 Measurement uncertainty	5
10 Documentation	5
11 Keywords	6
Annexes	6
Figure A1.1 Example of measurement of dose as function of average beam current I , conveying speed V and beam width W_b . Measured at an electron accelerator with beam energy 110 keV. $K = 216.57 \text{ (kGy} \cdot \text{m}^2 \text{) / (A} \cdot \text{s)}$	7
Figure A1.2 Example of beam width measurement (3 measurements and their average are shown). Beam width was measured on a low-energy accelerator installed in an electron beam tunnel for an aseptic filling line (3)	7
Figure A1.3 Example of beam width measurement at a low-energy electron accelerator facility for curing purpose	8
Figure A1.4 Calculated depth-dose distribution in water (specific density 1 g cm^{-3})	8
Figure A1.5 Calculated depth-dose distribution in water (specific density 1 g cm^{-3})	9
Figure A1.6 Methods for measurement of depth-dose distribution	9
Figure A1.7 Examples of measurements of depth-dose distributions at the same electron beam facility, but at different beam energies	10
Figure A2.1 Apparent dose measured with three dosimeters (18 μm RCD film dosimeter (1.12 g cm^{-3}); 50 μm RCD film dosimeter (1.15 g cm^{-3}); 130 μm alanine film dosimeter (1.36 g cm^{-3}) all calibrated by irradiation at a 10 MeV electron accelerator, and now irradiated at a 116 keV electron accelerator	11